

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU010522\
 Data File : VU046676.D
 Acq On : 05 Jan 2022 20:37
 Operator : SY/MD
 Sample : N1009-02
 Misc : 5.26g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGLF8

Quant Time: Jan 06 03:05:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM010522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jan 06 02:54:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	268851	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.427	117	267320	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	133151	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	125587	53.531	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 107.060%		
7) Chloroethane-d5	1.919	69	94986	54.501	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 109.000%		
11) 1,1-Dichloroethene-d2	2.555	63	147768	37.033	ug/L	-0.02
Spiked Amount	50.000	Range 60 - 125	Recovery	= 74.060%		
21) 2-Butanone-d5	4.658	46	190475	110.931	ug/L	0.02
Spiked Amount	100.000	Range 40 - 130	Recovery	= 110.930%		
24) Chloroform-d	5.063	84	224717	54.335	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 108.660%		
26) 1,2-Dichloroethane-d4	5.703	65	150255	57.285	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 114.560%		
32) Benzene-d6	5.726	84	465243	55.768	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 111.540%		
36) 1,2-Dichloropropane-d6	6.690	67	142003	55.787	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 111.580%		
41) Toluene-d8	7.899	98	434747	58.293	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 116.580%		
43) trans-1,3-Dichloroprop...	8.185	79	74592	60.826	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 121.660%		
47) 2-Hexanone-d5	8.674	63	167558	123.909	ug/L	0.04
Spiked Amount	100.000	Range 45 - 130	Recovery	= 123.910%		
56) 1,1,2,2-Tetrachloroeth...	10.771	84	220515	55.936	ug/L	0.01
Spiked Amount	50.000	Range 65 - 120	Recovery	= 111.880%		
66) 1,2-Dichlorobenzene-d4	12.198	152	177742	64.511	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 129.020%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU010522\
Data File : VU046676.D
Acq On : 05 Jan 2022 20:37
Operator : SY/MD
Sample : N1009-02
Misc : 5.26g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGLF8

Quant Time: Jan 06 03:05:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM010522WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jan 06 02:54:34 2022
Response via : Initial Calibration

