


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU010522\
 Data File : VU046657.D
 Acq On : 05 Jan 2022 10:02
 Operator : SY/MD
 Sample : BFB038
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB038

Integration File: LSCINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM010522WMA.M
 Title : VOC Analysis
 Last Update : Fri Jan 07 23:26:33 2022



AutoFind: Scans 2893, 2894, 2895; Background Corrected with Scan 2878

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.4	57909	PASS
75	95	30	80	52.2	174016	PASS
95	95	100	100	100.0	333120	PASS
96	95	5	9	6.7	22160	PASS
173	174	0.00	2	0.9	2288	PASS
174	95	50	120	76.8	255787	PASS
175	174	5	9	7.2	18531	PASS
176	174	95	101	96.4	246677	PASS
177	176	5	9	6.6	16307	PASS

m/z	Abundance
44.00	144.0

Instrument :
MSVOA_U
ClientSampleId :
BFB038

m/z	Abundance
36.00	2018.0
37.00	12348.0
38.00	10939.0
39.00	4532.0
40.10	188.0
44.00	1665.0
45.00	2260.0
46.00	188.0
47.00	3272.0
48.00	1615.0
49.00	11958.0
50.00	55672.0
51.00	17456.0
52.00	830.0
55.00	817.0
56.00	4613.0
57.00	7965.0
58.00	363.0
60.00	2954.0
61.00	14845.0
62.00	14784.0
63.00	11416.0
64.00	1294.0
65.00	925.0
67.00	787.0
68.00	32896.0
69.00	32832.0
70.00	2566.0
72.00	1891.0
73.00	14409.0
74.00	56464.0
75.00	168256.0
76.00	14227.0
77.00	1866.0
78.00	1241.0
78.90	10562.0
79.90	2936.0
81.00	10735.0
81.90	2164.0
82.90	201.0
85.90	257.0
87.00	13342.0
88.00	12300.0
90.90	1227.0
92.00	9298.0
93.00	14221.0
94.00	39200.0
95.10	319936.0
96.10	21776.0
97.00	745.0
102.90	177.0
103.90	1474.0
104.90	411.0
105.90	1539.0
106.90	388.0
109.90	167.0
110.90	309.0
111.90	286.0
113.00	283.0
114.90	312.0
115.90	1251.0
116.90	2276.0
117.90	1289.0
119.00	1905.0
123.90	145.0
125.80	127.0
127.90	1223.0
128.90	619.0
129.90	1152.0
130.90	407.0
134.90	666.0
136.90	599.0
138.90	129.0
140.00	232.0
141.00	3510.0
142.00	484.0
143.00	3913.0
143.90	219.0
144.90	573.0
145.90	550.0
146.90	221.0
148.00	853.0
148.90	224.0
149.90	267.0
151.90	142.0
152.90	285.0
153.90	181.0
155.00	636.0
155.90	136.0
157.00	503.0
158.90	421.0
160.90	385.0

Instrument :
MSVOA_U
ClientSampleId :
BFB038

171.00	207.0
172.00	1049.0
173.10	2127.0
174.00	240320.0
175.00	17752.0
176.00	230784.0
177.00	15457.0
178.00	379.0

Instrument :
MSVOA_U
ClientSampleId :
BFB038

m/z	Abundance
36.00	2164.0
37.00	12996.0
38.00	11592.0
39.00	4805.0
40.00	155.0
44.00	1795.0
45.00	2382.0
46.00	163.0
47.00	3329.0
48.00	1689.0
49.00	12686.0
50.00	59656.0
51.00	18584.0
52.00	884.0
55.00	870.0
56.00	5043.0
57.00	8551.0
58.00	441.0
60.00	3022.0
61.00	15704.0
62.00	15638.0
63.00	11854.0
64.00	1256.0
65.00	770.0
67.00	897.0
68.00	34296.0
69.00	34728.0
70.00	2665.0
72.00	1984.0
73.00	15676.0
74.00	59840.0
75.00	179008.0
76.00	15115.0
77.00	1960.0
78.00	1374.0
78.90	11067.0
79.90	3129.0
81.00	11637.0
81.90	2296.0
82.90	230.0
85.90	303.0
87.00	14069.0
88.00	13123.0
90.90	1282.0
92.00	9937.0
93.00	15434.0
94.00	41608.0
95.10	342784.0
96.10	22864.0
97.00	709.0
102.90	174.0
103.90	1508.0
104.90	454.0
105.90	1675.0
106.90	396.0
109.90	178.0
110.90	371.0
111.90	311.0
112.90	292.0
114.90	359.0
115.90	1341.0
117.00	2400.0
117.90	1454.0
119.00	2018.0
123.80	143.0
125.90	135.0
127.90	1291.0
128.90	684.0
130.00	1240.0
130.90	471.0
135.00	647.0
137.00	640.0
138.90	147.0
139.90	247.0
141.00	3664.0
141.90	480.0
143.00	3931.0
143.90	239.0
145.00	501.0
145.90	523.0
146.90	226.0
148.00	868.0
148.80	281.0
149.90	331.0
151.80	162.0
152.90	302.0
153.90	213.0
155.00	731.0
155.90	158.0
157.00	574.0
157.90	104.0
158.90	447.0

Instrument :
MSVOA_U
ClientSampleId :
BFB038

160.90	411.0
171.10	167.0
172.00	885.0
173.10	2361.0
174.00	262912.0
175.00	19208.0
176.00	253376.0
177.00	16808.0
178.00	363.0

Instrument :
MSVOA_U
ClientSampleId :
BFB038

m/z	Abundance
36.00	2126.0
37.00	12670.0
38.00	11425.0
39.00	4539.0
39.90	138.0
44.00	1759.0
45.00	2425.0
45.90	168.0
47.00	3098.0
48.00	1677.0
49.00	12324.0
50.00	58400.0
51.00	18056.0
52.00	826.0
55.00	802.0
56.00	4967.0
57.00	8543.0
58.00	456.0
60.00	2889.0
61.00	15282.0
62.00	15594.0
63.00	11370.0
64.00	1095.0
65.00	621.0
66.10	100.0
67.00	909.0
68.00	33056.0
69.00	33864.0
70.00	2550.0
72.00	1856.0
73.00	15528.0
74.00	58200.0
75.00	174784.0
76.00	14778.0
77.00	1876.0
78.00	1345.0
78.90	10390.0
79.90	3032.0
81.00	11332.0
81.90	2249.0
82.90	216.0
85.90	304.0
87.00	13803.0
88.00	12960.0
90.90	1239.0
92.00	9846.0
93.00	14985.0
94.00	40496.0
95.00	336640.0
96.10	21840.0
97.10	675.0
102.80	157.0
103.90	1456.0
104.90	461.0
105.90	1649.0
106.90	353.0
109.90	187.0
110.90	367.0
111.90	293.0
112.90	293.0
114.90	364.0
115.90	1335.0
117.00	2331.0
117.90	1494.0
119.00	1957.0
123.80	144.0
125.90	135.0
126.80	100.0
127.90	1277.0
128.90	666.0
130.00	1299.0
130.90	494.0
134.90	577.0
137.00	655.0
138.90	152.0
139.90	244.0
141.00	3636.0
141.90	446.0
143.00	3743.0
143.90	234.0
145.00	420.0
145.90	474.0
146.90	229.0
148.00	810.0
148.80	277.0
149.90	349.0
151.90	160.0
152.90	275.0
153.90	243.0
155.00	779.0
155.90	156.0
157.00	576.0

Instrument :
MSVOA_U
ClientSampleId :
BFB038

157.90	109.0
158.90	424.0
160.90	394.0
171.20	118.0
172.00	754.0
173.10	2376.0
174.00	264128.0
175.00	18632.0
176.00	255872.0
177.00	16656.0
178.00	360.0

Instrument :
MSVOA_U
ClientSampleId :
BFB038