

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU010622\
 Data File : VU046685.D
 Acq On : 06 Jan 2022 14:38
 Operator : SY/MD
 Sample : N1027-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0J16

Quant Time: Jan 07 03:15:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122121WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jan 07 03:14:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	72681	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	78445	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	38637	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	27687	4.836	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.800%
7) Chloroethane-d5	1.919	69	23035	5.110	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.200%
11) 1,1-Dichloroethene-d2	2.572	65	12135	5.027	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	100.600%
20) 2-Butanone-d5	4.655	46	93080	56.052	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.100%
24) Chloroform-d	5.070	84	54708	5.215	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.400%
26) 1,2-Dichloroethane-d4	5.706	65	32864	5.480	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.600%
32) Benzene-d6	5.729	84	106344	4.863	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.200%
36) 1,2-Dichloropropane-d6	6.694	67	33217	5.055	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.200%
41) Toluene-d8	7.899	98	90359	4.553	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.000%
43) trans-1,3-Dichloroprop...	8.179	79	14175	4.664	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.200%
46) 2-Hexanone-d5	8.639	63	86642	48.559	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.120%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	34152	5.064	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.200%
66) 1,2-Dichlorobenzene-d4	12.195	152	37265	5.236	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.800%
Target Compounds						
9) Trichlorofluoromethane	2.147	101	779	0.063	ug/L	98
12) 1,1-Dichloroethene	2.584	96	5761	0.901	ug/L #	1
19) 1,1-Dichloroethane	3.867	63	1651	0.131	ug/L #	80
22) cis-1,2-Dichloroethene	4.674	96	8442	1.132	ug/L	95
29) 1,1,1-Trichloroethane	5.321	97	5585	0.431	ug/L	96
34) Trichloroethene	6.546	95	43802	5.331	ug/L	98
37) 1,2-Dichloropropane	6.793	63	16277	2.047	ug/L	99
47) Tetrachloroethene	8.555	164	1740	0.299	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU010622\
Data File : VU046685.D
Acq On : 06 Jan 2022 14:38
Operator : SY/MD
Sample : N1027-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0J16

Quant Time: Jan 07 03:15:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122121WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jan 07 03:14:07 2022
Response via : Initial Calibration

