

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU010622\  
 Data File : VU046686.D  
 Acq On : 06 Jan 2022 15:02  
 Operator : SY/MD  
 Sample : N1027-08  
 Misc : 25.0mL/MSVOA\_U/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 C0J35

Quant Time: Jan 07 03:16:08 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMUTR122121WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Fri Jan 07 03:14:07 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon           | Response | Conc   | Units    | Dev(Min) |
|-------------------------------|--------|----------------|----------|--------|----------|----------|
| -----                         |        |                |          |        |          |          |
| Internal Standards            |        |                |          |        |          |          |
| 1) 1,4-Difluorobenzene        | 6.253  | 114            | 70888    | 5.000  | ug/L     | 0.00     |
| 28) Chlorobenzene-d5          | 9.420  | 117            | 76855    | 5.000  | ug/L     | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.812 | 152            | 37849    | 5.000  | ug/L     | 0.00     |
| System Monitoring Compounds   |        |                |          |        |          |          |
| 4) Vinyl Chloride-d3          | 1.604  | 65             | 27742    | 4.968  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 40 - 130 | Recovery | =      | 99.400%  |          |
| 7) Chloroethane-d5            | 1.919  | 69             | 23185    | 5.273  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 65 - 130 | Recovery | =      | 105.400% |          |
| 11) 1,1-Dichloroethene-d2     | 2.572  | 65             | 12227    | 5.193  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 60 - 125 | Recovery | =      | 103.800% |          |
| 20) 2-Butanone-d5             | 4.662  | 46             | 96659    | 59.680 | ug/L     | 0.00     |
| Spiked Amount                 | 50.000 | Range 40 - 130 | Recovery | =      | 119.360% |          |
| 24) Chloroform-d              | 5.067  | 84             | 53729    | 5.251  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 125 | Recovery | =      | 105.000% |          |
| 26) 1,2-Dichloroethane-d4     | 5.707  | 65             | 33308    | 5.695  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 130 | Recovery | =      | 113.800% |          |
| 32) Benzene-d6                | 5.729  | 84             | 108620   | 5.070  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 125 | Recovery | =      | 101.400% |          |
| 36) 1,2-Dichloropropane-d6    | 6.694  | 67             | 33008    | 5.127  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 60 - 140 | Recovery | =      | 102.600% |          |
| 41) Toluene-d8                | 7.899  | 98             | 89598    | 4.608  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 130 | Recovery | =      | 92.200%  |          |
| 43) trans-1,3-Dichloroprop... | 8.182  | 79             | 14191    | 4.766  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 55 - 130 | Recovery | =      | 95.400%  |          |
| 46) 2-Hexanone-d5             | 8.639  | 63             | 89402    | 51.142 | ug/L     | 0.00     |
| Spiked Amount                 | 50.000 | Range 45 - 130 | Recovery | =      | 102.280% |          |
| 56) 1,1,2,2-Tetrachloroeth... | 10.758 | 84             | 34750    | 5.259  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 65 - 120 | Recovery | =      | 105.200% |          |
| 66) 1,2-Dichlorobenzene-d4    | 12.195 | 152            | 38288    | 5.492  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 80 - 120 | Recovery | =      | 109.800% |          |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU010622\  
Data File : VU046686.D  
Acq On : 06 Jan 2022 15:02  
Operator : SY/MD  
Sample : N1027-08  
Misc : 25.0mL/MSVOA\_U/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
C0J35

Quant Time: Jan 07 03:16:08 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMUTR122121WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Fri Jan 07 03:14:07 2022  
Response via : Initial Calibration

