

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU010719\
 Data File : VU029059.D
 Acq On : 07 Jan 2019 15:36
 Operator : MD/SY
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTD05051

Quant Time: Jan 08 01:43:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM122018WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jan 08 01:37:07 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 200%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	0.408	0.375	8.1	90	0.00
3 T	Chloromethane	0.554	0.453	18.2	80	0.00
4 S	Vinyl Chloride-d3	0.401	0.393	2.0	92	0.00
5 T	Vinyl chloride	0.490	0.454	7.3	88	0.00
6 T	Bromomethane	0.209	0.201	3.8	98	0.00
7 S	Chloroethane-d5	0.310	0.314	-1.3	97	0.00
8 T	Chloroethane	0.280	0.257	8.2	87	0.00
9 T	Trichlorofluoromethane	0.501	0.487	2.8	92	0.00
10 T	1,1,2-Trichloro-1,2,2-trifl	0.324	0.318	1.9	94	0.00
11 S	1,1-Dichloroethene-d2	0.674	0.650	3.6	92	0.00
12 T	1,1-Dichloroethene	0.321	0.295	8.1	90	0.00
13 T	Acetone	0.286	0.275	3.8	91	0.00
14 T	Carbon disulfide	1.120	0.953	14.9	83	0.00
15 T	Methyl Acetate	0.446	0.422	5.4	91	0.00
16 T	Methylene chloride	0.395	0.384	2.8	95	0.00
17 T	trans-1,2-Dichloroethene	0.347	0.332	4.3	92	0.00
18 T	Methyl tert-butyl Ether	1.068	1.061	0.7	95	0.00
19 T	1,1-Dichloroethane	0.675	0.654	3.1	93	0.00
20 T	cis-1,2-Dichloroethene	0.393	0.384	2.3	94	0.00
21 S	2-Butanone-d5	0.281	0.297	-5.7	96	0.00
22 T	2-Butanone	0.349	0.339	2.9	91	0.00
23 T	Bromochloromethane	0.188	0.184	2.1	94	0.00
24 S	Chloroform-d	0.630	0.664	-5.4	98	0.00
25 T	Chloroform	0.643	0.638	0.8	96	0.00
26 S	1,2-Dichloroethane-d4	0.389	0.409	-5.1	99	0.00
27 T	1,2-Dichloroethane	0.478	0.470	1.7	94	0.00
28 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
29 T	Cyclohexane	0.665	0.606	8.9	87	0.00
30 T	1,1,1-Trichloroethane	0.529	0.526	0.6	96	0.00
31 T	Carbon tetrachloride	0.427	0.428	-0.2	97	0.00
32 S	Benzene-d6	1.413	1.463	-3.5	97	0.00
33 T	Benzene	1.592	1.548	2.8	93	0.00
34 T	Trichloroethene	0.384	0.362	5.7	92	0.00
35 T	Methylcyclohexane	0.640	0.617	3.6	91	0.00
36 S	1,2-Dichloropropane-d6	0.480	0.495	-3.1	98	0.00
37 T	1,2-Dichloropropane	0.445	0.437	1.8	95	0.00
38 T	Bromodichloromethane	0.493	0.495	-0.4	96	0.00
39 T	cis-1,3-Dichloropropene	0.634	0.628	0.9	95	0.00
40 T	4-Methyl-2-pentanone	0.598	0.595	0.5	95	0.00
41 S	Toluene-d8	1.280	1.323	-3.4	96	0.00
42 T	Toluene	1.614	1.609	0.3	95	0.00
43 S	trans-1,3-Dichloropropene-d	0.215	0.219	-1.9	94	0.00
44 T	trans-1,3-Dichloropropene	0.565	0.567	-0.4	94	0.00
45 T	1,1,2-Trichloroethane	0.380	0.384	-1.1	98	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU010719\
 Data File : VU029059.D
 Acq On : 07 Jan 2019 15:36
 Operator : MD/SY
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTD05051

Quant Time: Jan 08 01:43:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM122018WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jan 08 01:37:07 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 200%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Tetrachloroethene	0.296	0.293	1.0	97	0.00
47 S	2-Hexanone-d5	0.228	0.220	3.5	87	0.00
48 T	2-Hexanone	0.491	0.483	1.6	92	0.00
49 T	Dibromochloromethane	0.379	0.390	-2.9	99	0.00
50 T	1,2-Dibromoethane	0.396	0.399	-0.8	96	0.00
51 T	Chlorobenzene	1.017	1.029	-1.2	98	0.00
52 T	Ethylbenzene	1.734	1.704	1.7	94	0.00
53 T	m,p-Xylene	0.641	0.640	0.2	94	0.00
54 T	o-xylene	0.643	0.643	0.0	94	0.00
55 T	Styrene	1.088	1.105	-1.6	95	0.00
56 T	Isopropylbenzene	1.630	1.625	0.3	95	0.00
57 S	1,1,2,2-Tetrachloroethane-d	0.666	0.695	-4.4	100	0.00
58 T	1,1,2,2-Tetrachloroethane	0.691	0.697	-0.9	98	0.00
59	1,2,3-Trichloropropane	0.531	0.525	1.1	96	0.00
60 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
61 T	Bromoform	0.620	0.665	-7.3	101	0.00
62 T	1,3-Dichlorobenzene	1.588	1.604	-1.0	95	0.00
63 T	1,4-Dichlorobenzene	1.636	1.681	-2.8	98	0.00
64 S	1,2-Dichlorobenzene-d4	1.011	1.070	-5.8	98	0.00
65 T	1,2-Dichlorobenzene	1.657	1.756	-6.0	99	0.00
66 T	1,2-Dibromo-3-chloropropane	0.286	0.296	-3.5	98	0.00
67	1,3,5-Trichlorobenzene	1.171	1.169	0.2	93	0.00
68 T	1,2,4-trichlorobenzene	0.981	1.000	-1.9	94	0.00
69	Naphthalene	3.405	3.492	-2.6	94	0.00
70 T	1,2,3-Trichlorobenzene	1.057	1.084	-2.6	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0