

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU010720\
 Data File : VU036394.D
 Acq On : 07 Jan 2020 15:22
 Operator : JC/MD
 Sample : VU0107WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK33

Quant Time: Jan 08 04:16:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM122719WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jan 08 03:56:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	546145	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	526673	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	201802	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	172448	40.52	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.04%
7) Chloroethane-d5	1.93	69	155249	43.25	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.50%
11) 1,1-Dichloroethene-d2	2.59	63	261955	36.13	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.26%
21) 2-Butanone-d5	4.68	46	318514	100.51	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.51%
24) Chloroform-d	5.11	84	357788	48.17	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.34%
26) 1,2-Dichloroethane-d4	5.75	65	248939	50.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.10%
32) Benzene-d6	5.77	84	730828	49.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.68%
36) 1,2-Dichloropropane-d6	6.73	67	238051	48.41	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.82%
41) Toluene-d8	7.93	98	681682	48.43	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.86%
43) trans-1,3-Dichloropropene-	8.21	79	126332	47.34	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.68%
47) 2-Hexanone-d5	8.67	63	239738	94.53	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.53%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	344468	46.19	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.38%
64) 1,2-Dichlorobenzene-d4	12.22	152	200368	47.33	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.66%


Target Compounds					Qvalue
13) Acetone	2.66	43	5314	1.617 ug/L	82


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU010720\
Data File : VU036394.D
Acq On : 07 Jan 2020 15:22
Operator : JC/MD
Sample : VU0107WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK33

Quant Time: Jan 08 04:16:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM122719WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jan 08 03:56:17 2020
Response via : Initial Calibration





#13

Acetone

Concen: 1.617 ug/L

RT: 2.66 min Scan# 411

Delta R.T. 0.00 min

Lab File: VU036394.D

Acq: 07 Jan 2020 15:22

Instrument :

MSVOA_U

ClientSampleId :

VBLK33

Tgt Ion: 43 Resp: 5314

Ion Ratio Lower Upper

43 100

58 23.2 0.0 66.2

