

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU010722\
 Data File : VU046695.D
 Acq On : 07 Jan 2022 13:46
 Operator : SY/MD
 Sample : VU0107WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK057

Quant Time: Jan 07 23:27:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM010522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jan 07 23:26:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	189849	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	189854	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.815	152	89536	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	59477	35.902	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	71.800%
7) Chloroethane-d5	1.919	69	49350	40.099	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	80.200%
11) 1,1-Dichloroethene-d2	2.571	63	84683	30.054	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	60.100%
21) 2-Butanone-d5	4.649	46	113068	93.252	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.250%
24) Chloroform-d	5.070	84	121580	41.630	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.260%
26) 1,2-Dichloroethane-d4	5.706	65	82210	44.385	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.780%
32) Benzene-d6	5.732	84	238856	40.314	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.620%
36) 1,2-Dichloropropane-d6	6.693	67	75607	41.823	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	83.640%
41) Toluene-d8	7.902	98	212331	40.087	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	80.180%
43) trans-1,3-Dichloroprop...	8.182	79	36724	42.166	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.340%
47) 2-Hexanone-d5	8.639	63	96117	100.081	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.080%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	126716	45.258	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.520%
66) 1,2-Dichlorobenzene-d4	12.195	152	89081	48.081	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.160%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU010722\
Data File : VU046695.D
Acq On : 07 Jan 2022 13:46
Operator : SY/MD
Sample : VU0107WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK057

Quant Time: Jan 07 23:27:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM010522WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jan 07 23:26:33 2022
Response via : Initial Calibration

