


Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU010821\
 Data File : VU041863.D
 Acq On : 07 Jan 2021 09:13
 Operator : SY/MD
 Sample : BFB35
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB35

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SFAMUSIM120820.M
 Title : TRACE VOA SOM01.0
 Last Update : Fri Jan 08 15:27:30 2021



AutoFind: Scans 2889, 2890, 2891; Background Corrected with Scan 2874

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.7	22122	PASS
75	95	30	80	48.4	67994	PASS
95	95	100	100	100.0	140522	PASS
96	95	5	9	6.3	8906	PASS
173	174	0.00	2	0.4	544	PASS
174	95	50	120	86.2	121074	PASS
175	174	5	9	7.3	8881	PASS
176	174	95	101	95.4	115509	PASS
177	176	5	9	6.5	7501	PASS

m/z	Abundance
36.00	979.0
37.00	5476.0
38.10	4273.0
39.00	2365.0
40.00	230.0
44.00	679.0
45.10	1002.0
45.90	251.0
47.10	1558.0
48.00	802.0
49.10	5226.0
50.10	23008.0
51.10	6321.0
52.00	310.0
53.10	252.0
55.00	297.0
56.10	1605.0
57.10	3299.0
58.00	155.0
60.10	1143.0
61.10	6292.0
62.10	6538.0
63.10	5077.0
64.10	911.0
65.10	1339.0
66.10	269.0
67.10	444.0
68.10	12951.0
69.10	13733.0
70.00	936.0
72.00	881.0
73.10	5416.0
74.10	22696.0
75.10	69440.0
76.10	5903.0
77.00	978.0
78.00	468.0
79.00	4120.0
80.00	1199.0
81.00	4310.0
82.00	1118.0
86.00	185.0
87.10	6109.0
88.10	5782.0
91.00	535.0
92.10	3589.0
93.10	5938.0
94.10	15510.0
95.10	143104.0
96.10	9180.0
97.10	290.0
103.90	597.0
105.00	180.0
105.90	412.0
106.90	130.0
110.80	145.0
115.00	153.0
116.00	461.0
117.00	1168.0
118.00	467.0
118.90	951.0
128.00	515.0
129.00	294.0
130.00	377.0
131.00	250.0
135.10	158.0
137.00	181.0
141.10	1133.0
142.10	143.0
143.00	1882.0
144.00	179.0
145.10	626.0
145.90	174.0
147.90	245.0
150.00	130.0
155.00	350.0
157.10	275.0
159.00	111.0
161.10	103.0
171.10	244.0
172.10	1852.0
173.10	1632.0
174.10	121224.0
175.10	9078.0
176.10	115792.0
177.10	7856.0
178.00	222.0

Instrument :
MSVOA_U
ClientSampleId :
BFB35

Scan 2874 (10.581 min): VU041863.D
BFB35

m/z	Abundance
43.90	122.0

Instrument :
MSVOA_U
ClientSampleId :
BFB35

m/z	Abundance
36.00	867.0
37.00	4768.0
38.10	4059.0
39.10	1984.0
39.90	124.0
44.00	662.0
45.00	819.0
45.80	166.0
47.10	1279.0
48.00	627.0
49.10	4754.0
50.10	20512.0
51.10	6634.0
52.10	332.0
53.00	170.0
55.00	357.0
56.00	1663.0
57.10	2817.0
58.00	135.0
60.10	1060.0
61.10	5389.0
62.10	5698.0
63.10	4857.0
64.10	734.0
65.10	1393.0
66.00	198.0
67.10	367.0
68.10	12225.0
69.10	12159.0
70.00	872.0
72.00	798.0
73.10	5112.0
74.10	21144.0
75.10	64056.0
76.10	5061.0
77.00	795.0
77.90	501.0
79.00	3374.0
80.00	1048.0
81.00	3899.0
82.00	1088.0
86.10	173.0
87.10	5933.0
88.10	5401.0
91.00	421.0
92.10	3443.0
93.10	5479.0
94.10	14368.0
95.10	133632.0
96.10	8282.0
97.10	388.0
104.00	375.0
105.00	210.0
106.00	497.0
106.90	157.0
115.00	137.0
116.00	382.0
117.00	950.0
118.00	494.0
119.00	783.0
128.00	531.0
129.00	212.0
130.00	333.0
131.00	271.0
135.10	171.0
137.00	169.0
141.10	1049.0
142.10	207.0
143.00	1697.0
144.00	103.0
145.10	522.0
146.10	214.0
148.00	286.0
149.10	154.0
150.10	106.0
155.10	301.0
157.10	269.0
159.00	221.0
161.00	120.0
171.10	198.0
172.10	1759.0
174.10	117080.0
175.10	8354.0
176.10	111528.0
177.10	6968.0
178.00	174.0

Instrument :
MSVOA_U
ClientSampleId :
BFB35

m/z	Abundance
36.00	971.0
37.00	5302.0
38.10	4303.0
39.00	2279.0
40.00	187.0
44.00	690.0
45.10	986.0
45.90	239.0
47.10	1385.0
48.00	722.0
49.10	5256.0
50.10	22848.0
51.10	6931.0
52.00	350.0
53.10	216.0
55.10	370.0
56.00	1723.0
57.10	3185.0
58.00	146.0
60.10	1165.0
61.10	6054.0
62.10	6337.0
63.10	5248.0
64.10	868.0
65.10	1395.0
66.00	239.0
67.10	431.0
68.10	13405.0
69.10	13657.0
70.00	992.0
72.00	932.0
73.10	5409.0
74.10	22792.0
75.10	70488.0
76.10	5758.0
77.00	908.0
78.00	502.0
79.00	3923.0
80.00	1148.0
81.00	4274.0
82.00	1180.0
86.10	164.0
87.10	6401.0
88.10	5889.0
91.00	490.0
92.10	3626.0
93.10	5929.0
94.10	15749.0
95.10	144832.0
96.10	9256.0
97.10	337.0
103.90	487.0
105.00	203.0
105.90	463.0
106.90	181.0
110.80	135.0
115.00	140.0
116.00	422.0
117.00	1113.0
118.00	490.0
119.00	910.0
128.00	554.0
129.00	269.0
129.90	389.0
131.00	268.0
135.20	162.0
137.00	186.0
141.10	1120.0
142.10	177.0
143.00	1854.0
144.00	169.0
145.10	580.0
146.00	187.0
148.00	278.0
149.10	134.0
150.00	133.0
155.00	335.0
157.10	300.0
159.00	200.0
161.00	118.0
171.10	254.0
172.10	1890.0
174.10	124920.0
175.10	9212.0
176.10	119208.0
177.10	7680.0
178.00	204.0

Instrument :
MSVOA_U
ClientSampleId :
BFB35