

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU010822\
 Data File : VU046704.D
 Acq On : 08 Jan 2022 14:25
 Operator : SY/MD
 Sample : N1077-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFXW1

Quant Time: Jan 10 00:40:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122121WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jan 10 00:25:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.285	114	80389	5.000	ug/L	0.04
28) Chlorobenzene-d5	9.427	117	87186	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	53096	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.610	65	24081	3.803	ug/L	0.02
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.000%
7) Chloroethane-d5	1.938	69	20984	4.208	ug/L	0.03
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.200%
11) 1,1-Dichloroethene-d2	2.584	65	10396	3.894	ug/L	0.02
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.800%
20) 2-Butanone-d5	4.790	46	70644	38.463	ug/L	0.14
Spiked Amount	50.000	Range	40 - 130	Recovery	=	76.920%
24) Chloroform-d	5.096	84	54108	4.664	ug/L	0.04
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.200%
26) 1,2-Dichloroethane-d4	5.755	65	32256	4.863	ug/L	0.05
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.200%
32) Benzene-d6	5.771	84	101499	4.176	ug/L	0.05
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.600%
36) 1,2-Dichloropropane-d6	6.726	67	32598	4.464	ug/L	0.04
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.200%
41) Toluene-d8	7.915	98	86484	3.921	ug/L	0.02
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.400%
43) trans-1,3-Dichloroprop...	8.198	79	13472	3.989	ug/L	0.02
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.800%
46) 2-Hexanone-d5	8.661	63	77948	39.306	ug/L	0.02
Spiked Amount	50.000	Range	45 - 130	Recovery	=	78.620%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	36135	4.821	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.400%
66) 1,2-Dichlorobenzene-d4	12.195	152	41850	4.279	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	85.600%
Target Compounds						Qvalue
15) Methyl Acetate	2.970	43	1762	0.507	ug/L #	51
35) Methylcyclohexane	6.723	83	8416	0.654	ug/L #	24
37) 1,2-Dichloropropane	6.723	63	3485	0.394	ug/L #	88
42) Toluene	7.986	91	12146	0.342	ug/L	100
48) 2-Hexanone	8.803	43	287585	58.055	ug/L #	17
61) 1,2,3-Trichloropropane	11.816	75	6093	0.691	ug/L #	60


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU010822\
Data File : VU046704.D
Acq On : 08 Jan 2022 14:25
Operator : SY/MD
Sample : N1077-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFXW1

Quant Time: Jan 10 00:40:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR122121WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Jan 10 00:25:50 2022
Response via : Initial Calibration









#15
Methyl Acetate
Concen: 0.507 ug/L
RT: 2.970 min Scan# 505
Delta R.T. 0.006 min
Lab File: VU046704.D
Acq: 08 Jan 2022 14:25




Instrument :
MSVOA_U
ClientSampleId :
BFXW1


Tgt Ion: 43 Resp: 1762
Ion Ratio Lower Upper
43 100
74 0.0 19.1 28.7#



#35
Methylcyclohexane
Concen: 0.654 ug/L
RT: 6.723 min Scan# 1674
Delta R.T. -0.042 min
Lab File: VU046704.D
Acq: 08 Jan 2022 14:25

Tgt Ion: 83 Resp: 8416
Ion Ratio Lower Upper
83 100
55 9.6 61.9 92.9#
98 0.0 39.0 58.6#





#37

1,2-Dichloropropane

Concen: 0.394 ug/L

RT: 6.723 min Scan# 1

Delta R.T. -0.067 min

Lab File: VU046704.D

Acq: 08 Jan 2022 14:25

Instrument :

MSVOA_U

ClientSampleId :





BFXW1

Tgt Ion: 63 Resp: 3485

Ion Ratio Lower Upper

63 100

112 0.0 3.2 4.8#



#42

Toluene

Concen: 0.342 ug/L

RT: 7.986 min Scan# 2067

Delta R.T. 0.016 min

Lab File: VU046704.D




Acq: 08 Jan 2022 14:25


Tgt Ion: 91 Resp: 12146

Ion Ratio Lower Upper

91 100

92 59.0 41.3 76.7









#48
2-Hexanone
Concen: 58.055 ug/L
RT: 8.803 min Scan# 21
Delta R.T. 0.113 min
Lab File: VU046704.D
Acq: 08 Jan 2022 14:25

Instrument :
MSVOA_U
ClientSampleId :
BFXW1

Tgt Ion: 43 Resp: 287585
Ion Ratio Lower Upper
43 100
58 19.4 46.7 70.1#
57 115.7 16.3 24.5#
100 0.9 10.5 15.7#



#61
1,2,3-Trichloropropane
Concen: 0.691 ug/L
RT: 11.816 min Scan# 3258
Delta R.T. 0.990 min
Lab File: VU046704.D
Acq: 08 Jan 2022 14:25

Tgt Ion: 75 Resp: 6093
Ion Ratio Lower Upper
75 100
110 0.0 29.9 44.9#
77 39.5 25.5 38.3#

