

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU010824\
 Data File : VU057421.D
 Acq On : 09 Jan 2024 07:10
 Operator : MD/SY
 Sample : P1058-17
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :

Quant Time: Jan 09 07:37:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM010824WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jan 09 06:09:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	258760	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	251447	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	105258	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	70434	39.548	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	79.100%		
7) Chloroethane-d5	1.914	69	55962	40.605	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	81.200%		
11) 1,1-Dichloroethene-d2	2.560	63	142133	37.652	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	75.300%		
21) 2-Butanone-d5	4.631	46	86406	100.125	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	100.130%		
24) Chloroform-d	5.055	84	176197	41.837	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	83.680%		
26) 1,2-Dichloroethane-d4	5.695	65	124243	44.656	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	89.320%		
32) Benzene-d6	5.721	84	314028	46.525	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	93.040%		
36) 1,2-Dichloropropane-d6	6.685	67	89763	47.119	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	94.240%		
41) Toluene-d8	7.891	98	274055	43.328	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	86.660%		
43) trans-1,3-Dichloroprop...	8.174	79	50497	44.409	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	88.820%		
47) 2-Hexanone-d5	8.644	63	59595	95.694	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	95.690%		
56) 1,1,2,2-Tetrachloroeth...	10.753	84	125414	44.053	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	88.100%		
66) 1,2-Dichlorobenzene-d4	12.190	152	97286	47.256	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	94.520%		
Target Compounds						Qvalue
5) Vinyl chloride	1.605	62	3867	1.950	ug/L #	73
12) 1,1-Dichloroethene	2.573	96	47496	27.089	ug/L #	83
13) Acetone	2.647	43	29707	28.268	ug/L	73
19) 1,1-Dichloroethane	3.866	63	10364	3.121	ug/L	94
20) cis-1,2-Dichloroethene	4.660	96	13531	7.052	ug/L	96
30) 1,1,1-Trichloroethane	5.309	97	12041	3.112	ug/L	97
34) Trichloroethene	6.537	95	17528	8.496	ug/L	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU010824\
Data File : VU057421.D
Acq On : 09 Jan 2024 07:10
Operator : MD/SY
Sample : P1058-17
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :


Quant Time: Jan 09 07:37:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM010824WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Jan 09 06:09:25 2024
Response via : Initial Calibration








#5
 Vinyl chloride
 Concen: 1.950 ug/L
 RT: 1.605 min Scan# 91
 Delta R.T. -0.000 min
 Lab File: VU057421.D
 Acq: 09 Jan 2024 07:10


Instrument :
 MSVOA_U
 ClientSampleId :





Tgt Ion: 62 Resp: 3867
 Ion Ratio Lower Upper
 62 100
 64 48.8 23.3 43.3#




#12
 1,1-Dichloroethene
 Concen: 27.089 ug/L
 RT: 2.573 min Scan# 399
 Delta R.T. 0.003 min
 Lab File: VU057421.D
 Acq: 09 Jan 2024 07:10



Tgt Ion: 96 Resp: 47496
 Ion Ratio Lower Upper
 96 100
 61 182.7 128.3 238.3
 63 175.6 90.4 168.0#





#13

Acetone

Concen: 28.268 ug/L

RT: 2.647 min Scan# 41

Delta R.T. -0.000 min

Lab File: VU057421.D

Acq: 09 Jan 2024 07:10

Instrument :

MSVOA_U





ClientSampleId :

Tgt Ion: 43 Resp: 29707

Ion Ratio Lower Upper

43 100

58 17.2 0.0 65.2



#19

1,1-Dichloroethane

Concen: 3.121 ug/L

RT: 3.866 min Scan# 801

Delta R.T. 0.006 min

Lab File: VU057421.D

Acq: 09 Jan 2024 07:10




Tgt Ion: 63 Resp: 10364


Ion Ratio Lower Upper

63 100

65 36.0 22.4 41.6

83 17.2 11.3 21.1









#20
 cis-1,2-Dichloroethene
 Concen: 7.052 ug/L
 RT: 4.660 min Scan# 1048
 Delta R.T. 0.003 min
 Lab File: VU057421.D
 Acq: 09 Jan 2024 07:10




Instrument :
 MSVOA_U
 ClientSampleId :


Tgt Ion: 96 Resp: 13531
 Ion Ratio Lower Upper
 96 100
 61 124.3 88.0 163.4
 98 70.3 43.4 80.6



#30
 1,1,1-Trichloroethane
 Concen: 3.112 ug/L
 RT: 5.309 min Scan# 1250
 Delta R.T. -0.000 min
 Lab File: VU057421.D
 Acq: 09 Jan 2024 07:10


Tgt Ion: 97 Resp: 12041
 Ion Ratio Lower Upper
 97 100
 99 68.0 52.1 78.1
 61 41.8 31.9 47.9





#34
 Trichloroethene
 Concen: 8.496 ug/L
 RT: 6.537 min Scan# 1
 Delta R.T. 0.003 min
 Lab File: VU057421.D
 Acq: 09 Jan 2024 07:10

Instrument :
 MSVOA_U
 ClientSampleId :



Tgt Ion: 95 Resp: 17528

Ion	Ratio	Lower	Upper
95	100		
97	64.3	46.2	85.8
132	91.7	64.3	119.5
130	95.4	67.2	124.8

