

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU010919\
 Data File : VU029088.D
 Acq On : 09 Jan 2019 14:36
 Operator : MD/SY
 Sample : PB116248ZHE#03
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB116248ZHE#03

Quant Time: Jan 10 04:57:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 04 02:22:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	111451	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	190187	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	177975	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	69763	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	74291	53.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.84%	
35) Dibromofluoromethane	4.88	113	65681	53.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.56%	
50) Toluene-d8	7.56	98	244744	50.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.58%	
62) 4-Bromofluorobenzene	10.30	95	77181	44.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.24%	
Target Compounds						
16) Acetone	2.32	43	5733	6.605	ug/l	98
18) Methyl Acetate	2.62	43	2035	1.008	ug/l #	90
20) Methylene Chloride	2.70	84	2552	1.732	ug/l	86
43) Isopropyl Acetate	5.55	43	9624	3.007	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU010919\
Data File : VU029088.D
Acq On : 09 Jan 2019 14:36
Operator : MD/SY
Sample : PB116248ZHE#03
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB116248ZHE#03

Quant Time: Jan 10 04:57:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
Quant Title : SW846 8260
QLast Update : Fri Jan 04 02:22:53 2019
Response via : Initial Calibration

