

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU010919\
 Data File : VU029089.D
 Acq On : 09 Jan 2019 15:00
 Operator : MD/SY
 Sample : PB116248ZHE#04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB116248ZHE#04

Quant Time: Jan 10 04:59:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 04 02:22:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	109026	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	189964	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	179762	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	72928	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	75227	55.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.64%	
35) Dibromofluoromethane	4.88	113	65567	53.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.50%	
50) Toluene-d8	7.56	98	249985	51.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.84%	
62) 4-Bromofluorobenzene	10.30	95	78391	44.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.74%	
Target Compounds						
16) Acetone	2.32	43	5452	6.421	ug/l	97
20) Methylene Chloride	2.70	84	2064	1.432	ug/l	94
43) Isopropyl Acetate	5.55	43	9566	2.992	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU010919\
Data File : VU029089.D
Acq On : 09 Jan 2019 15:00
Operator : MD/SY
Sample : PB116248ZHE#04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB116248ZHE#04

Quant Time: Jan 10 04:59:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
Quant Title : SW846 8260
QLast Update : Fri Jan 04 02:22:53 2019
Response via : Initial Calibration

