

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU010919\
 Data File : VU029093.D
 Acq On : 09 Jan 2019 16:37
 Operator : MD/SY
 Sample : PB116248ZHE#08
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB116248ZHE#08

Quant Time: Jan 10 05:35:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 04 02:22:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	107735	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	188695	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	177533	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	68377	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	74831	56.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.38%	
35) Dibromofluoromethane	4.88	113	64347	52.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.22%	
50) Toluene-d8	7.56	98	244934	50.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.44%	
62) 4-Bromofluorobenzene	10.30	95	74040	42.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.32%	
Target Compounds						
16) Acetone	2.32	43	5234	6.238	ug/l	98
20) Methylene Chloride	2.70	84	2004	1.407	ug/l	95
43) Isopropyl Acetate	5.56	43	8427	2.654	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU010919\
Data File : VU029093.D
Acq On : 09 Jan 2019 16:37
Operator : MD/SY
Sample : PB116248ZHE#08
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB116248ZHE#08

Quant Time: Jan 10 05:35:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
Quant Title : SW846 8260
QLast Update : Fri Jan 04 02:22:53 2019
Response via : Initial Calibration

