

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU010919\
 Data File : VU029100.D
 Acq On : 09 Jan 2019 19:27
 Operator : MD/SY
 Sample : PB116248ZHE#15
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB116248ZHE#15

Quant Time: Jan 10 05:47:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 04 02:22:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	107136	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	186570	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	178347	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	70588	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	75263	56.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.66%	
35) Dibromofluoromethane	4.88	113	64279	53.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.30%	
50) Toluene-d8	7.56	98	243553	51.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.02%	
62) 4-Bromofluorobenzene	10.31	95	76206	44.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.82%	
Target Compounds						
16) Acetone	2.32	43	5618	6.734	ug/l	97
20) Methylene Chloride	2.70	84	2226	1.571	ug/l	95
43) Isopropyl Acetate	5.55	43	8066	2.569	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU010919\
Data File : VU029100.D
Acq On : 09 Jan 2019 19:27
Operator : MD/SY
Sample : PB116248ZHE#15
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB116248ZHE#15

Quant Time: Jan 10 05:47:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
Quant Title : SW846 8260
QLast Update : Fri Jan 04 02:22:53 2019
Response via : Initial Calibration

