

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU010921\
 Data File : VU041875.D
 Acq On : 07 Jan 2021 16:17
 Operator : SY/MD
 Sample : M1026-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFXK2

Quant Time: Jan 08 05:40:48 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jan 08 04:58:19 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	134100	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	129413	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	54671	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	43240	4.13	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.60%
7) Chloroethane-d5	1.92	69	45869	5.65	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	113.00%
11) 1,1-Dichloroethene-d2	2.58	63	78945	4.09	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.80%
20) 2-Butanone-d5	4.66	46	185451	61.58	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	123.16%
24) Chloroform-d	5.08	84	104020	5.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	115.80%
26) 1,2-Dichloroethane-d4	5.72	65	58425	5.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	114.60%
32) Benzene-d6	5.74	84	188637	5.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.80%
36) 1,2-Dichloropropane-d6	6.70	67	59708	5.14	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.80%
41) Toluene-d8	7.90	98	144976	4.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.60%
43) trans-1,3-Dichloropropene-	8.18	79	23214	4.44	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.80%
46) 2-Hexanone-d5	8.64	63	142498	49.90	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.80%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	56121	5.32	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.40%
64) 1,2-Dichlorobenzene-d4	12.19	152	55626	5.72	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.40%

Target Compounds

					Qvalue
2) Dichlorodifluoromethane	1.39	85	1105	0.097 ug/L #	74
3) Chloromethane	1.52	50	889	0.081 ug/L	92
9) Trichlorofluoromethane	2.15	101	5878554	381.089 ug/L	99
10) 1,1,2-Trichloro-1,2,2-trif	2.59	101	99475	10.783 ug/L	98
12) 1,1-Dichloroethene	2.59	96	24952	2.743 ug/L #	75
13) Acetone	2.68	43	2678	1.398 ug/L	96
16) Methylene chloride	3.06	84	2573	0.237 ug/L	93
17) Methyl tert-butyl Ether	3.38	73	183402	7.556 ug/L	99
19) 1,1-Dichloroethane	3.89	63	125831	7.155 ug/L	97
22) cis-1,2-Dichloroethene	4.68	96	4582	0.441 ug/L	76
25) Chloroform	5.10	83	11001	0.621 ug/L	88
29) 1,1,1-Trichloroethane	5.34	97	11420	0.768 ug/L	97
31) Carbon tetrachloride	5.53	117	1408	0.110 ug/L	82
34) Trichloroethene	6.55	95	50958	4.900 ug/L	97
47) Tetrachloroethene	8.56	164	1946	0.267 ug/L	93

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU010921\
Data File : VU041875.D
Acq On : 07 Jan 2021 16:17
Operator : SY/MD
Sample : M1026-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFXK2

Quant Time: Jan 08 05:40:48 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jan 08 04:58:19 2021
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#)=qualifier out of range (m)=manual integration (+)=signals summed						

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU010921\
Data File : VU041875.D
Acq On : 07 Jan 2021 16:17
Operator : SY/MD
Sample : M1026-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFXK2

Quant Time: Jan 08 05:40:48 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jan 08 04:58:19 2021
Response via : Initial Calibration





#2

Dichlorodifluoromethane

Concen: 0.097 ug/L

RT: 1.39 min Scan# 15

Delta R.T. 0.00 min

Lab File: VU041875.D

Acq: 07 Jan 2021 16:17

Instrument :

MSVOA_U

ClientSampleId :


BFXK2

Tgt Ion: 85 Resp: 1105

Ion Ratio Lower Upper




85 100

87 46.5 25.8 38.6#



Abundance Ion 85.00 (84.70 to 85.70): VU0

Ion 87.00 (86.70 to 87.70): VU0



#3

Chloromethane

Concen: 0.081 ug/L

RT: 1.52 min Scan# 57

Delta R.T. -0.00 min

Lab File: VU041875.D


Acq: 07 Jan 2021 16:17

Tgt Ion: 50 Resp: 889

Ion Ratio Lower Upper



50 100


52 29.8 24.1 44.8



Abundance Ion 50.05 (49.75 to 50.75): VU0

Ion 52.05 (51.75 to 52.75): VU0









#9

Trichlorofluoromethane
 Concen: 381.089 ug/L
 RT: 2.15 min Scan# 251
 Delta R.T. 0.00 min
 Lab File: VU041875.D
 Acq: 07 Jan 2021 16:17

Instrument :
 MSVOA_U
 ClientSampleId :
 BFXK2




Tgt Ion:101 Resp: 5878554
 Ion Ratio Lower Upper
 101 100
 103 65.0 51.1 76.7




#10

1,1,2-Trichloro-1,2,2-trifluoroethane
 Concen: 10.783 ug/L
 RT: 2.59 min Scan# 389
 Delta R.T. 0.00 min
 Lab File: VU041875.D
 Acq: 07 Jan 2021 16:17

Tgt Ion:101 Resp: 99475
 Ion Ratio Lower Upper
 101 100
 85 42.7 33.6 50.4
 151 72.9 56.2 84.4





#12

1,1-Dichloroethene

Concen: 2.743 ug/L

RT: 2.59 min Scan# 389

Delta R.T. 0.00 min

Lab File: VU041875.D

Acq: 07 Jan 2021 16:17

Instrument :

MSVOA_U

ClientSampleId :

BFXK2


Tgt Ion: 96 Resp: 24952

Ion Ratio Lower Upper

96 100

61 158.2 123.8 230.0




63 162.4 81.0 150.4#



Abundance Ion 96.00 (95.70 to 96.70): VU041869.D (-375) (-)

Ion 61.05 (60.75 to 61.75): VU041869.D (-375) (-)

Ion 63.00 (62.70 to 63.70): VU041869.D (-375) (-)



#13

Acetone

Concen: 1.398 ug/L

RT: 2.68 min Scan# 417

Delta R.T. 0.02 min

Lab File: VU041875.D


Acq: 07 Jan 2021 16:17

Tgt Ion: 43 Resp: 2678

Ion Ratio Lower Upper



43 100


58 34.0 0.0 73.2



Abundance Ion 43.05 (42.75 to 43.75): VU041869.D (-395) (-)

Ion 58.05 (57.75 to 58.75): VU041869.D (-395) (-)





#16

Methylene chloride

Concen: 0.237 ug/L

RT: 3.06 min Scan# 535

Delta R.T. 0.00 min

Lab File: VU041875.D

Acq: 07 Jan 2021 16:17

Instrument :

MSVOA_U

ClientSampleId :

BFXK2


Tgt Ion: 84 Resp: 2573

Ion Ratio Lower Upper

84 100

86 58.3 45.1 83.7




49 111.7 83.1 154.3



Abundance Ion 84.05 (83.75 to 84.75): VU0

Ion 86.00 (85.70 to 86.70): VU0

Ion 49.10 (48.80 to 49.80): VU0



#17

Methyl tert-butyl Ether

Concen: 7.556 ug/L

RT: 3.38 min Scan# 635

Delta R.T. 0.01 min

Lab File: VU041875.D

Acq: 07 Jan 2021 16:17


Tgt Ion: 73 Resp: 183402

Ion Ratio Lower Upper

73 100

43 18.7 14.7 22.1



57 22.4 18.2 27.4




Abundance Ion 73.10 (72.80 to 73.80): VU0

Ion 43.05 (42.75 to 43.75): VU0

Ion 57.10 (56.80 to 57.80): VU0









#19
 1,1-Dichloroethane
 Concen: 7.155 ug/L
 RT: 3.89 min Scan# 792
 Delta R.T. 0.01 min
 Lab File: VU041875.D
 Acq: 07 Jan 2021 16:17

Instrument :
 MSVOA_U
 ClientSampleId :
 BFXK2

Tgt Ion: 63 Resp: 125831
 Ion Ratio Lower Upper
 63 100
 65 32.5 21.3 39.6
 83 13.3 9.3 17.3




Abundance Ion 63.10 (62.80 to 63.80): VU0
 Ion 65.10 (64.80 to 65.80): VU0
 Ion 83.05 (82.75 to 83.75): VU0






#22
 cis-1,2-Dichloroethene
 Concen: 0.441 ug/L
 RT: 4.68 min Scan# 1040
 Delta R.T. 0.01 min
 Lab File: VU041875.D
 Acq: 07 Jan 2021 16:17

Tgt Ion: 96 Resp: 4582
 Ion Ratio Lower Upper
 96 100
 61 154.6 88.8 165.0
 68 0.0 0.0 0.0



Abundance Ion 96.00 (95.70 to 96.70): VU0
 Ion 61.05 (60.75 to 61.75): VU0
 Ion 68.15 (67.85 to 68.85): VU0









#25
Chloroform
Concen: 0.621 ug/L
RT: 5.10 min Scan# 1170
Delta R.T. 0.00 min
Lab File: VU041875.D
Acq: 07 Jan 2021 16:17




Instrument :
MSVOA_U
ClientSampleId :
BFXK2


Tgt Ion: 83 Resp: 11001
Ion Ratio Lower Upper
83 100
85 53.6 44.2 82.0



#29
1,1,1-Trichloroethane
Concen: 0.768 ug/L
RT: 5.34 min Scan# 1243
Delta R.T. 0.01 min
Lab File: VU041875.D
Acq: 07 Jan 2021 16:17

Tgt Ion: 97 Resp: 11420
Ion Ratio Lower Upper
97 100
99 67.9 51.3 76.9
61 46.3 36.5 54.7





#31

Carbon tetrachloride

Concen: 0.110 ug/L

RT: 5.53 min Scan# 1304

Delta R.T. -0.00 min

Lab File: VU041875.D

Acq: 07 Jan 2021 16:17

Instrument :

MSVOA_U

ClientSampleId :


BFXK2

Tgt Ion: 117 Resp: 1408

Ion Ratio Lower Upper




117 100

119 112.8 76.2 114.2



Abundance Ion 116.90 (116.60 to 117.60): V

Ion 118.90 (118.60 to 119.60): V



#34

Trichloroethene

Concen: 4.900 ug/L

RT: 6.55 min Scan# 1621

Delta R.T. 0.00 min

Lab File: VU041875.D

Acq: 07 Jan 2021 16:17

Tgt Ion: 95 Resp: 50958


Ion Ratio Lower Upper

95 100

97 65.7 43.9 81.5

132 94.4 63.1 117.3

130 97.7 67.4 125.2






Abundance Ion 95.00 (94.70 to 95.70): VU0

Ion 96.95 (96.65 to 97.65): VU0

Ion 131.95 (131.65 to 132.65): VU0

Ion 129.95 (129.65 to 130.65): VU0





#47

Tetrachloroethene

Concen: 0.267 ug/L

RT: 8.56 min Scan# 2245

Delta R.T. 0.00 min

Lab File: VU041875.D


Acq: 07 Jan 2021 16:17

Instrument :

MSVOA_U

ClientSampleId :

BFXK2



Tgt Ion:164 Resp: 1946


Ion Ratio Lower Upper

164 100

129 98.2 72.2 134.0

131 88.8 66.9 124.3

166 117.5 89.6 166.4



Abundance Ion 163.90 (163.60 to 164.60): V

Ion 128.95 (128.65 to 129.65): V

Ion 130.95 (130.65 to 131.65): V

Ion 165.90 (165.60 to 166.60): V

