


Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU010921\
 Data File : VU041891.D
 Acq On : 08 Jan 2021 09:14
 Operator : SY/MD
 Sample : BFB31
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB31

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
 Title : TRACE VOA SOM01.0
 Last Update : Mon Jan 11 16:09:52 2021



AutoFind: Scans 2888, 2889, 2890; Background Corrected with Scan 2874

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.5	23024	PASS
75	95	30	80	48.5	72149	PASS
95	95	100	100	100.0	148821	PASS
96	95	5	9	7.1	10500	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	84.9	126280	PASS
175	174	5	9	7.9	9951	PASS
176	174	95	101	98.0	123696	PASS
177	176	5	9	6.2	7707	PASS

Scan 2874 (10.581 min): VU041891.D
BFB31

m/z	Abundance
43.90	241.0

Instrument :
MSVOA_U
ClientSampleId :
BFB31

m/z	Abundance
36.00	866.0
37.10	5065.0
38.10	4482.0
39.10	1841.0
40.10	192.0
44.00	797.0
45.10	895.0
46.10	229.0
47.00	1041.0
48.00	700.0
49.10	4907.0
50.10	22128.0
51.10	7371.0
52.00	219.0
53.00	106.0
54.90	297.0
56.10	1784.0
57.10	3103.0
58.00	126.0
60.10	1031.0
61.10	5787.0
62.10	6247.0
63.10	5397.0
64.10	941.0
65.10	1240.0
66.10	186.0
67.20	338.0
68.10	11870.0
69.10	12718.0
70.00	904.0
72.10	676.0
73.10	6189.0
74.10	21152.0
75.10	68592.0
76.10	6281.0
77.10	853.0
77.90	663.0
79.00	3723.0
80.00	1008.0
81.00	4322.0
82.00	927.0
86.10	215.0
87.00	6311.0
88.00	5712.0
91.00	455.0
92.10	3853.0
93.10	6319.0
94.10	16520.0
95.10	139968.0
96.10	10036.0
97.10	333.0
104.00	524.0
105.00	177.0
106.00	581.0
107.00	181.0
110.00	100.0
113.00	104.0
114.90	145.0
116.00	541.0
117.00	850.0
118.00	577.0
119.00	878.0
127.90	433.0
129.10	172.0
130.10	498.0
130.90	147.0
135.10	187.0
136.90	239.0
139.80	151.0
141.10	1238.0
142.20	181.0
143.00	1719.0
144.00	135.0
145.00	624.0
146.00	249.0
148.10	283.0
150.10	184.0
152.20	102.0
155.00	302.0
157.00	172.0
161.00	149.0
171.00	282.0
172.10	1780.0
174.10	115224.0
175.10	9309.0
176.10	112904.0
177.10	7418.0
178.10	196.0

Instrument :
MSVOA_U
ClientSampleId :
BFB31

m/z	Abundance
36.00	952.0
37.00	5236.0
38.10	4928.0
39.10	1991.0
40.10	230.0
44.00	787.0
45.10	998.0
46.10	216.0
47.00	1302.0
48.00	770.0
49.10	5288.0
50.10	23792.0
51.10	7893.0
52.00	364.0
55.10	323.0
56.10	1993.0
57.10	3304.0
57.90	158.0
60.00	1092.0
61.10	6263.0
62.10	6544.0
63.10	5574.0
64.10	960.0
65.10	1334.0
66.10	155.0
67.20	364.0
68.10	13267.0
69.10	14191.0
70.00	1099.0
72.00	714.0
73.10	6744.0
74.10	22880.0
75.10	74336.0
76.10	6719.0
77.10	934.0
77.90	594.0
79.00	4019.0
80.00	1049.0
81.00	4322.0
82.00	1028.0
86.10	199.0
87.00	6821.0
88.00	6119.0
91.00	534.0
92.10	4298.0
93.10	6861.0
94.10	17904.0
95.10	153856.0
96.10	10909.0
97.10	300.0
104.00	610.0
105.10	181.0
106.00	614.0
107.00	198.0
110.00	111.0
110.90	103.0
112.10	100.0
113.00	137.0
115.00	137.0
116.00	594.0
117.00	923.0
118.00	600.0
119.00	952.0
127.90	509.0
129.10	209.0
130.00	559.0
130.90	164.0
135.10	145.0
137.00	280.0
139.80	112.0
141.10	1274.0
142.20	166.0
143.00	1822.0
143.90	121.0
145.00	639.0
146.00	229.0
148.00	321.0
150.10	131.0
153.10	105.0
153.90	114.0
155.00	349.0
157.00	228.0
159.10	143.0
161.20	156.0
171.00	287.0
172.10	1997.0
174.10	130304.0
175.10	10243.0
176.10	127648.0
177.10	7902.0
178.10	197.0

Instrument :
MSVOA_U
ClientSampleId :
BFB31

m/z	Abundance
36.00	863.0
37.00	4968.0
38.10	4744.0
39.10	1949.0
40.00	225.0
42.90	134.0
44.00	747.0
45.10	1057.0
46.10	208.0
47.00	1320.0
48.00	780.0
49.10	5218.0
50.10	23152.0
51.10	7827.0
52.10	403.0
55.10	377.0
56.10	2017.0
57.00	3178.0
58.00	197.0
60.00	1132.0
61.10	6273.0
62.10	6316.0
63.10	5104.0
64.10	928.0
65.10	1374.0
66.10	138.0
67.20	351.0
68.10	13603.0
69.10	14168.0
70.00	1055.0
71.90	756.0
73.10	6525.0
74.10	22592.0
75.10	73520.0
76.10	6455.0
77.10	987.0
78.00	561.0
79.00	3919.0
80.00	990.0
81.00	4017.0
82.00	1010.0
82.90	116.0
86.20	158.0
87.00	6617.0
88.00	6069.0
91.00	532.0
92.10	4174.0
93.10	6684.0
94.10	17208.0
95.10	152640.0
96.10	10555.0
97.10	227.0
104.00	613.0
105.10	120.0
105.90	543.0
107.00	162.0
110.90	128.0
112.10	115.0
112.90	155.0
114.90	157.0
116.00	574.0
117.00	935.0
118.10	563.0
119.00	936.0
127.90	513.0
129.10	231.0
130.00	584.0
130.90	157.0
135.10	135.0
137.00	241.0
141.10	1152.0
142.10	156.0
143.00	1642.0
143.90	114.0
145.00	600.0
146.00	143.0
148.00	296.0
150.00	118.0
153.10	114.0
153.90	103.0
155.10	355.0
157.00	256.0
159.10	176.0
161.20	161.0
171.00	289.0
172.10	1956.0
174.10	133312.0
175.10	10302.0
176.10	130536.0
177.10	7803.0
178.00	204.0

Instrument :
MSVOA_U
ClientSampleId :
BFB31