

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU011019\
 Data File : VU029125.D
 Acq On : 10 Jan 2019 18:18
 Operator : MD/SY
 Sample : K1123-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CPGP-01

Quant Time: Jan 12 02:13:33 2019
 Quant Method : Z:\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 04 02:22:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	128516	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	226018	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	206121	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	86712	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	85761	53.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.96%	
35) Dibromofluoromethane	4.88	113	73701	50.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.60%	
50) Toluene-d8	7.56	98	282775	48.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.78%	
62) 4-Bromofluorobenzene	10.30	95	91950	44.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.46%	
Target Compounds						
11) Tert butyl alcohol	2.83	59	20380	50.467	ug/l	98
16) Acetone	2.32	43	2148	2.146	ug/l	95
19) Methyl tert-butyl Ether	3.00	73	3809	0.813	ug/l	97
65) Chlorobenzene	9.12	112	12559	2.718	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU011019\
Data File : VU029125.D
Acq On : 10 Jan 2019 18:18
Operator : MD/SY
Sample : K1123-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CPGP-01

Quant Time: Jan 12 02:13:33 2019
Quant Method : Z:\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
Quant Title : SW846 8260
QLast Update : Fri Jan 04 02:22:53 2019
Response via : Initial Calibration

