

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU011019\
 Data File : VU029128.D
 Acq On : 10 Jan 2019 19:31
 Operator : MD/SY
 Sample : K1123-03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CPGP-03

Quant Time: Jan 11 04:13:43 2019
 Quant Method : Z:\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 04 02:22:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	126262	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	216681	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	198209	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	80510	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	83378	53.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.84%	
35) Dibromofluoromethane	4.88	113	72899	51.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.80%	
50) Toluene-d8	7.56	98	272995	49.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.46%	
62) 4-Bromofluorobenzene	10.30	95	87467	43.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.78%	
Target Compounds						
16) Acetone	2.32	43	1971	2.005	ug/l	99
19) Methyl tert-butyl Ether	3.00	73	2945	0.640	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU011019\
Data File : VU029128.D
Acq On : 10 Jan 2019 19:31
Operator : MD/SY
Sample : K1123-03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CPGP-03

Quant Time: Jan 11 04:13:43 2019
Quant Method : Z:\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
Quant Title : SW846 8260
QLast Update : Fri Jan 04 02:22:53 2019
Response via : Initial Calibration

