

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011022\
 Data File : VU046716.D
 Acq On : 10 Jan 2022 15:51
 Operator : SY/MD
 Sample : N1053-17ME
 Misc : 6.06g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGLJ4ME

Quant Time: Jan 11 01:20:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM010522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jan 07 23:26:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	259441	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	256114	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.815	152	134868	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	70559	31.166	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	62.340%
7) Chloroethane-d5	1.919	69	60742	36.116	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	72.240%
11) 1,1-Dichloroethene-d2	2.568	63	98326	25.536	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	51.080%#
21) 2-Butanone-d5	4.655	46	161433	97.427	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.430%
24) Chloroform-d	5.067	84	152755	38.275	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	76.540%
26) 1,2-Dichloroethane-d4	5.706	65	101655	40.162	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.320%
32) Benzene-d6	5.732	84	301323	37.700	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	75.400%
36) 1,2-Dichloropropane-d6	6.693	67	95506	39.162	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	78.320%
41) Toluene-d8	7.902	98	267011	37.369	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	74.740%#
43) trans-1,3-Dichloroprop...	8.185	79	47454	40.390	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.780%
47) 2-Hexanone-d5	8.652	63	133387	102.956	ug/L	0.01
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.960%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	172944	45.789	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.580%
66) 1,2-Dichlorobenzene-d4	12.198	152	124463	44.599	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011022\
Data File : VU046716.D
Acq On : 10 Jan 2022 15:51
Operator : SY/MD
Sample : N1053-17ME
Misc : 6.06g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGLJ4ME

Quant Time: Jan 11 01:20:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM010522WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jan 07 23:26:33 2022
Response via : Initial Calibration

