

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011022\
 Data File : VU046712.D
 Acq On : 10 Jan 2022 12:30
 Operator : SY/MD
 Sample : VU0110WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK059

Quant Time: Jan 10 14:49:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM010522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jan 07 23:26:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	183145	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	177624	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	87171	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	66738	41.759	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.520%
7) Chloroethane-d5	1.919	69	52366	44.107	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	88.220%
11) 1,1-Dichloroethene-d2	2.572	63	90858	33.426	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.860%
21) 2-Butanone-d5	4.652	46	104085	88.985	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	88.990%
24) Chloroform-d	5.067	84	117259	41.620	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.240%
26) 1,2-Dichloroethane-d4	5.707	65	77651	43.458	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.920%
32) Benzene-d6	5.732	84	244246	44.062	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.120%
36) 1,2-Dichloropropane-d6	6.694	67	73451	43.428	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	86.860%
41) Toluene-d8	7.903	98	216769	43.743	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.480%
43) trans-1,3-Dichloroprop...	8.182	79	36157	44.373	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.740%
47) 2-Hexanone-d5	8.639	63	84165	93.670	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.670%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	112929	43.111	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.220%
66) 1,2-Dichlorobenzene-d4	12.195	152	87225	48.357	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.720%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011022\
Data File : VU046712.D
Acq On : 10 Jan 2022 12:30
Operator : SY/MD
Sample : VU0110WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK059

Quant Time: Jan 10 14:49:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM010522WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jan 07 23:26:33 2022
Response via : Initial Calibration

