

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU011121\
 Data File : VU041942.D
 Acq On : 12 Jan 2021 00:12
 Operator : SY/MD
 Sample : M1026-20
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFXM1

Quant Time: Jan 12 07:11:42 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR011121WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jan 12 05:50:11 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	110608	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	105782	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	45413	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	44538	4.77	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	95.40%
7) Chloroethane-d5	1.92	69	43526	5.04	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.80%
11) 1,1-Dichloroethene-d2	2.58	63	66650	3.77	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.40%
20) 2-Butanone-d5	4.65	46	142505	55.00	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.00%
24) Chloroform-d	5.08	84	81547	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.00%
26) 1,2-Dichloroethane-d4	5.72	65	48083	5.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.40%
32) Benzene-d6	5.74	84	159802	5.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.20%
36) 1,2-Dichloropropane-d6	6.70	67	54145	5.51	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.20%
41) Toluene-d8	7.90	98	133879	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.40%
43) trans-1,3-Dichloropropene-	8.19	79	20837	5.08	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	101.60%
46) 2-Hexanone-d5	8.64	63	119056	57.42	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	114.84%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	47225	5.43	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	108.60%
64) 1,2-Dichlorobenzene-d4	12.19	152	49347	5.75	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	115.00%

Target Compounds

					Ovalue
3) Chloromethane	1.52	50	925	0.079 ug/L	96
16) Methylene chloride	3.06	84	2275	0.210 ug/L	89
22) cis-1,2-Dichloroethene	4.69	96	1291	0.144 ug/L	98
25) Chloroform	5.11	83	2321	0.142 ug/L	89
29) 1,1,1-Trichloroethane	5.34	97	2377	0.172 ug/L	96
34) Trichloroethene	6.55	95	48833	5.457 ug/L	97
47) Tetrachloroethene	8.56	164	30607	4.915 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU011121\
Data File : VU041942.D
Acq On : 12 Jan 2021 00:12
Operator : SY/MD
Sample : M1026-20
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFXM1

Quant Time: Jan 12 07:11:42 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR011121WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jan 12 05:50:11 2021
Response via : Initial Calibration

