

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011122\
 Data File : VU046727.D
 Acq On : 11 Jan 2022 15:49
 Operator : SY/MD
 Sample : N1074-21ME
 Misc : 6.22g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGLM4ME

Quant Time: Jan 12 04:51:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM010522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jan 12 04:49:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	217090	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	217241	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.815	152	109304	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	78912	41.656	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.320%
7) Chloroethane-d5	1.919	69	70932	50.403	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	100.800%
11) 1,1-Dichloroethene-d2	2.568	63	105020	32.595	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.180%
21) 2-Butanone-d5	4.652	46	267437	192.889	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	192.890%#
24) Chloroform-d	5.067	84	196689	58.897	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	117.800%
26) 1,2-Dichloroethane-d4	5.706	65	132539	62.579	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	125.160%#
32) Benzene-d6	5.729	84	346183	51.063	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.120%
36) 1,2-Dichloropropane-d6	6.693	67	120215	58.115	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	116.220%
41) Toluene-d8	7.902	98	285585	47.120	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.240%
43) trans-1,3-Dichloroprop...	8.182	79	60840	61.049	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	122.100%
47) 2-Hexanone-d5	8.655	63	218369	198.710	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	198.710%#
56) 1,1,2,2-Tetrachloroeth...	10.764	84	267808	83.593	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	167.180%#
66) 1,2-Dichlorobenzene-d4	12.198	152	166531	73.629	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	147.260%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011122\
Data File : VU046727.D
Acq On : 11 Jan 2022 15:49
Operator : SY/MD
Sample : N1074-21ME
Misc : 6.22g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGLM4ME

Quant Time: Jan 12 04:51:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM010522WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jan 12 04:49:52 2022
Response via : Initial Calibration

