

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011122\
 Data File : VU046724.D
 Acq On : 11 Jan 2022 13:12
 Operator : SY/MD
 Sample : N1074-12
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK002

Quant Time: Jan 12 04:50:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM010522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jan 12 04:49:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	170826	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	170511	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.813	152	79368	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	63622	42.680	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	85.360%
7) Chloroethane-d5	1.919	69	51209	46.243	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.480%
11) 1,1-Dichloroethene-d2	2.572	63	86099	33.960	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.920%
21) 2-Butanone-d5	4.649	46	102471	93.923	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.920%
24) Chloroform-d	5.067	84	117100	44.561	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.120%
26) 1,2-Dichloroethane-d4	5.703	65	75770	45.464	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.920%
32) Benzene-d6	5.729	84	235879	44.328	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.660%
36) 1,2-Dichloropropane-d6	6.691	67	70858	43.642	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	87.280%
41) Toluene-d8	7.900	98	209587	44.058	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.120%
43) trans-1,3-Dichloroprop...	8.179	79	34625	44.266	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.540%
47) 2-Hexanone-d5	8.636	63	81532	94.525	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.520%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	112449	44.719	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.440%
66) 1,2-Dichlorobenzene-d4	12.192	152	84715	51.583	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.160%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011122\
Data File : VU046724.D
Acq On : 11 Jan 2022 13:12
Operator : SY/MD
Sample : N1074-12
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK002

Quant Time: Jan 12 04:50:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM010522WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jan 12 04:49:52 2022
Response via : Initial Calibration

