

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011122\
 Data File : VU046728.D
 Acq On : 11 Jan 2022 16:12
 Operator : SY/MD
 Sample : N1074-03ME
 Misc : 6.02g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGLK2ME

Quant Time: Jan 12 04:51:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM010522WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jan 12 04:49:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	171657	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	174506	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	85743	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	42633	28.461	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	56.920%#
7) Chloroethane-d5	1.919	69	39648	35.630	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	71.260%
11) 1,1-Dichloroethene-d2	2.568	63	57557	22.592	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	45.180%#
21) 2-Butanone-d5	4.649	46	134782	122.941	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	122.940%
24) Chloroform-d	5.067	84	109096	41.314	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.620%
26) 1,2-Dichloroethane-d4	5.703	65	74020	44.199	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.400%
32) Benzene-d6	5.729	84	188020	34.525	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	69.040%#
36) 1,2-Dichloropropane-d6	6.693	67	65873	39.643	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	79.280%
41) Toluene-d8	7.899	98	155562	31.953	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	63.900%#
43) trans-1,3-Dichloroprop...	8.182	79	32499	40.597	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.200%
47) 2-Hexanone-d5	8.648	63	108517	122.930	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	122.930%
56) 1,1,2,2-Tetrachloroeth...	10.761	84	143905	55.918	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	111.840%
66) 1,2-Dichlorobenzene-d4	12.195	152	90854	51.208	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.420%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011122\
Data File : VU046728.D
Acq On : 11 Jan 2022 16:12
Operator : SY/MD
Sample : N1074-03ME
Misc : 6.02g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGLK2ME

Quant Time: Jan 12 04:51:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM010522WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jan 12 04:49:52 2022
Response via : Initial Calibration

