

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011123\
 Data File : VU052654.D
 Acq On : 11 Jan 2023 16:58
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ICSVU011123

Quant Time: Jan 12 03:54:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U011123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 12 03:51:10 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	103	0.00
2 T	Dichlorodifluoromethane	0.630	0.618	1.9	99	0.00
3 P	Chloromethane	0.917	0.866	5.6	106	0.00
4 C	Vinyl Chloride	0.790	0.780	1.3#	105	0.00
5 T	Bromomethane	0.459	0.446	2.8	106	0.00
6 T	Chloroethane	0.474	0.455	4.0	106	0.00
7 T	Trichlorofluoromethane	0.992	0.943	4.9	100	0.00
8 T	Diethyl Ether	0.408	0.406	0.5	106	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.593	0.562	5.2	98	0.00
10 T	Methyl Iodide	0.614	0.649	-5.7	109	0.00
11 T	Tert butyl alcohol	0.157	0.167	-6.4	118	0.00
12 CM	1,1-Dichloroethene	0.584	0.571	2.2#	103	0.00
13 T	Acrolein	0.189	0.189	0.0	115	0.00
14 T	Allyl chloride	1.150	1.197	-4.1	109	0.00
15 T	Acrylonitrile	0.422	0.445	-5.5	113	0.00
16 T	Acetone	0.410	0.361	12.0	93	0.00
17 T	Carbon Disulfide	1.851	1.786	3.5	102	0.00
18 T	Methyl Acetate	1.194	1.261	-5.6	112	0.00
19 T	Methyl tert-butyl Ether	2.019	2.101	-4.1	107	0.00
20 T	Methylene Chloride	0.863	0.714	17.3	104	0.00
21 T	trans-1,2-Dichloroethene	0.632	0.638	-0.9	105	0.00
22 T	Diisopropyl ether	2.211	2.336	-5.7	107	0.00
23 T	Vinyl Acetate	1.722	1.873	-8.8	109	0.00
24 P	1,1-Dichloroethane	1.301	1.296	0.4	105	0.00
25 T	2-Butanone	0.562	0.588	-4.6	108	0.00
26 T	2,2-Dichloropropane	1.014	0.991	2.3	103	0.00
27 T	cis-1,2-Dichloroethene	0.722	0.715	1.0	105	0.00
28 T	Bromochloromethane	0.598	0.615	-2.8	107	0.00
29 T	Tetrahydrofuran	0.341	0.376	-10.3	114	0.00
30 C	Chloroform	1.323	1.215	8.2#	104	0.00
31 T	Cyclohexane	1.232	1.215	1.4	103	0.00
32 T	1,1,1-Trichloroethane	1.036	1.010	2.5	103	0.00
33 S	1,2-Dichloroethane-d4	0.799	0.808	-1.1	109	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	109	0.00
35 S	Dibromofluoromethane	0.344	0.346	-0.6	110	0.00
36 T	1,1-Dichloropropene	0.542	0.522	3.7	102	0.00
37 T	Ethyl Acetate	0.687	0.641	6.7	112	0.00
38 T	Carbon Tetrachloride	0.503	0.479	4.8	102	0.00
39 T	Methylcyclohexane	0.623	0.623	0.0	101	0.00
40 TM	Benzene	1.602	1.581	1.3	106	0.00
41 T	Methacrylonitrile	0.322	0.340	-5.6	110	0.00
42 TM	1,2-Dichloroethane	0.576	0.553	4.0	104	0.00
43 T	Isopropyl Acetate	0.901	0.945	-4.9	111	0.00
44 TM	Trichloroethene	0.379	0.356	6.1	102	0.00
45 C	1,2-Dichloropropane	0.444	0.426	4.1#	104	0.00
46 T	Dibromomethane	0.278	0.271	2.5	107	0.00
47 T	Bromodichloromethane	0.571	0.541	5.3	104	0.00
48 T	Methyl methacrylate	0.419	0.453	-8.1	115	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011123\
 Data File : VU052654.D
 Acq On : 11 Jan 2023 16:58
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ICVVU011123

Quant Time: Jan 12 03:54:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U011123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 12 03:51:10 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.009	-12.5	112	0.00
50 S	Toluene-d8	1.322	1.351	-2.2	111	0.00
51 T	4-Methyl-2-Pentanone	0.589	0.636	-8.0	112	0.00
52 CM	Toluene	0.942	0.945	-0.3#	104	0.00
53 T	t-1,3-Dichloropropene	0.602	0.622	-3.3	108	0.00
54 T	cis-1,3-Dichloropropene	0.656	0.673	-2.6	107	0.00
55 T	1,1,2-Trichloroethane	0.389	0.372	4.4	105	0.00
56 T	Ethyl methacrylate	0.573	0.627	-9.4	109	0.00
57 T	1,3-Dichloropropane	0.694	0.694	0.0	106	0.00
58 T	2-Chloroethyl Vinyl ether	0.021	0.023	-9.5	121	0.00
59 T	2-Hexanone	0.450	0.487	-8.2	110	0.00
60 T	Dibromochloromethane	0.404	0.394	2.5	106	0.00
61 T	1,2-Dibromoethane	0.394	0.394	0.0	106	0.00
62 S	4-Bromofluorobenzene	0.479	0.496	-3.5	110	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	109	0.00
64 T	Tetrachloroethene	0.324	0.313	3.4	101	0.00
65 PM	Chlorobenzene	1.065	1.033	3.0	105	0.00
66 T	1,1,1,2-Tetrachloroethane	0.385	0.376	2.3	104	0.00
67 C	Ethyl Benzene	1.847	1.914	-3.6#	105	0.00
68 T	m/p-Xylenes	0.696	0.734	-5.5	104	0.00
69 T	o-Xylene	0.662	0.710	-7.3	106	0.00
70 T	Styrene	1.110	1.206	-8.6	106	0.00
71 P	Bromoform	0.323	0.315	2.5	107	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
73 T	Isopropylbenzene	3.422	3.674	-7.4	105	0.00
74 T	N-amyl acetate	1.564	1.755	-12.2	112	0.00
75 P	1,1,2,2-Tetrachloroethane	1.444	1.402	2.9	109	0.00
76 T	1,2,3-Trichloropropane	1.584	1.571	0.8	108	0.00
77 T	Bromobenzene	0.852	0.842	1.2	106	0.00
78 T	n-propylbenzene	4.163	4.501	-8.1	103	0.00
79 T	2-Chlorotoluene	2.546	2.649	-4.0	104	0.00
80 T	1,3,5-Trimethylbenzene	2.881	3.109	-7.9	104	0.00
81 T	trans-1,4-Dichloro-2-butene	0.415	0.443	-6.7	111	0.00
82 T	4-Chlorotoluene	2.904	3.057	-5.3	102	0.00
83 T	tert-Butylbenzene	2.871	3.029	-5.5	104	0.00
84 T	1,2,4-Trimethylbenzene	2.840	3.099	-9.1	104	0.00
85 T	sec-Butylbenzene	3.621	3.911	-8.0	103	0.00
86 T	p-Isopropyltoluene	2.966	3.261	-9.9	102	0.00
87 T	1,3-Dichlorobenzene	1.642	1.635	0.4	104	0.00
88 T	1,4-Dichlorobenzene	1.685	1.645	2.4	102	0.00
89 T	n-Butylbenzene	2.699	2.895	-7.3	101	0.00
90 T	Hexachloroethane	0.582	0.571	1.9	101	0.00
91 T	1,2-Dichlorobenzene	1.632	1.635	-0.2	103	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.309	0.314	-1.6	110	0.00
93 T	1,2,4-Trichlorobenzene	0.958	1.017	-6.2	103	0.00
94 T	Hexachlorobutadiene	0.505	0.490	3.0	100	0.00
95 T	Naphthalene	2.869	3.333	-16.2	109	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011123\
Data File : VU052654.D
Acq On : 11 Jan 2023 16:58
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ICVVU011123

Quant Time: Jan 12 03:54:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U011123W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 12 03:51:10 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.957	1.001	-4.6	103	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6