

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU011219\
 Data File : VU029148.D
 Acq On : 11 Jan 2019 17:54
 Operator : JC/SP
 Sample : K1136-03
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-107

Quant Time: Jan 14 11:03:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 04 02:22:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	170099	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	295175	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	272969	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	111433	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	118768	56.48	ug/l	0.00
Spiked Amount						
			Recovery	=	112.96%	
35) Dibromofluoromethane	4.88	113	100382	52.46	ug/l	0.00
Spiked Amount						
			Recovery	=	104.92%	
50) Toluene-d8	7.57	98	375660	49.73	ug/l	0.00
Spiked Amount						
			Recovery	=	99.46%	
62) 4-Bromofluorobenzene	10.30	95	117673	43.35	ug/l	0.00
Spiked Amount						
			Recovery	=	86.70%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

