

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011223\
 Data File : VU052657.D
 Acq On : 12 Jan 2023 11:16
 Operator : JC/MD
 Sample : VU0112WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0112WBL01

Quant Time: Jan 13 01:42:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U011123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 13 01:24:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.369	168	217354	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.247	114	386486	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.414	117	350851	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.809	152	150729	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.697	65	187050	53.865	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.720%
35) Dibromofluoromethane	5.285	113	136919	51.433	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.860%
50) Toluene-d8	7.896	98	492523	48.191	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.380%
62) 4-Bromofluorobenzene	10.629	95	160922	43.446	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011223\
Data File : VU052657.D
Acq On : 12 Jan 2023 11:16
Operator : JC/MD
Sample : VU0112WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0112WBL01

Quant Time: Jan 13 01:42:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U011123W.M
Quant Title : SW846 8260
QLast Update : Fri Jan 13 01:24:13 2023
Response via : Initial Calibration

