

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011321\
 Data File : VU042005.D
 Acq On : 14 Jan 2021 01:28
 Operator : SY/MD
 Sample : M1065-10
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFXP0

Quant Time: Jan 15 02:49:11 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR011321WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jan 15 02:46:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.259	114	106623	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	104638	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.809	152	49794	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	43258	5.37	ug/L	0.00
Spiked Amount 5.000	Range 40 - 130		Recovery = 107.40%			
7) Chloroethane-d5	1.919	69	43435	5.70	ug/L	0.00
Spiked Amount 5.000	Range 65 - 130		Recovery = 114.00%			
11) 1,1-Dichloroethene-d2	2.575	63	62868	3.88	ug/L	0.00
Spiked Amount 5.000	Range 60 - 125		Recovery = 77.60%			
20) 2-Butanone-d5	4.649	46	167060	61.87	ug/L	0.00
Spiked Amount 50.000	Range 40 - 130		Recovery = 123.74%			
24) Chloroform-d	5.079	84	92757	5.63	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery = 112.60%			
26) 1,2-Dichloroethane-d4	5.716	65	50998	5.68	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery = 113.60%			
32) Benzene-d6	5.742	84	166763	4.93	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery = 98.60%			
36) 1,2-Dichloropropane-d6	6.700	67	49525	4.46	ug/L	0.00
Spiked Amount 5.000	Range 60 - 140		Recovery = 89.20%			
41) Toluene-d8	7.906	98	123444	4.08	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery = 81.60%			
43) trans-1,3-Dichloroprop...	8.185	79	20108	4.51	ug/L	0.00
Spiked Amount 5.000	Range 55 - 130		Recovery = 90.20%			
46) 2-Hexanone-d5	8.639	63	111559	48.80	ug/L	0.00
Spiked Amount 50.000	Range 45 - 130		Recovery = 97.60%			
57) 1,1,2,2-Tetrachloroeth...	10.754	84	47087	5.57	ug/L	0.00
Spiked Amount 5.000	Range 65 - 120		Recovery = 111.40%			
64) 1,2-Dichlorobenzene-d4	12.188	152	46135	4.92	ug/L	0.00
Spiked Amount 5.000	Range 80 - 120		Recovery = 98.40%			
Target Compounds						Qvalue
9) Trichlorofluoromethane	2.144	101	4221	0.27	ug/L	89
13) Acetone	2.652	43	3492	2.07	ug/L	86
17) Methyl tert-butyl Ether	3.379	73	4940	0.23	ug/L #	83
22) cis-1,2-Dichloroethene	4.681	96	7926	0.89	ug/L	86
25) Chloroform	5.102	83	5374	0.32	ug/L	90
34) Trichloroethene	6.552	95	117848	12.17	ug/L	98
47) Tetrachloroethene	8.558	164	24896	3.97	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011321\
Data File : VU042005.D
Acq On : 14 Jan 2021 01:28
Operator : SY/MD
Sample : M1065-10
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFXP0

Quant Time: Jan 15 02:49:11 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR011321WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jan 15 02:46:12 2021
Response via : Initial Calibration

