

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011321\
 Data File : VU042014.D
 Acq On : 14 Jan 2021 05:04
 Operator : SY/MD
 Sample : M1129-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFT61

Quant Time: Jan 15 02:52:02 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR011321WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jan 15 02:46:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.343	114	82903	5.00	ug/L	0.08
28) Chlorobenzene-d5	9.439	117	68931	5.00	ug/L	0.02
60) 1,4-Dichlorobenzene-d4	11.816	152	23861	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.652	65	40200	6.42	ug/L	0.05
Spiked Amount	5.000	Range	40 - 130	Recovery	=	128.40%
7) Chloroethane-d5	1.986	69	39686	6.70	ug/L	0.07
Spiked Amount	5.000	Range	65 - 130	Recovery	=	134.00%#
11) 1,1-Dichloroethene-d2	2.639	63	64661	5.13	ug/L	0.06
Spiked Amount	5.000	Range	60 - 125	Recovery	=	102.60%
20) 2-Butanone-d5	4.809	46	116384	55.44	ug/L	0.15
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.88%
24) Chloroform-d	5.137	84	94965	7.41	ug/L	0.06
Spiked Amount	5.000	Range	70 - 125	Recovery	=	148.20%#
26) 1,2-Dichloroethane-d4	5.813	65	38914	5.57	ug/L	0.10
Spiked Amount	5.000	Range	70 - 130	Recovery	=	111.40%
32) Benzene-d6	5.829	84	125388	5.62	ug/L	0.09
Spiked Amount	5.000	Range	70 - 125	Recovery	=	112.40%
36) 1,2-Dichloropropane-d6	6.774	67	42815	5.86	ug/L	0.07
Spiked Amount	5.000	Range	60 - 140	Recovery	=	117.20%
41) Toluene-d8	7.944	98	84992	4.27	ug/L	0.04
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.40%
43) trans-1,3-Dichloroprop...	8.221	79	15612	5.32	ug/L	0.04
Spiked Amount	5.000	Range	55 - 130	Recovery	=	106.40%
46) 2-Hexanone-d5	8.677	63	73211	48.62	ug/L	0.04
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.24%
57) 1,1,2,2-Tetrachloroeth...	10.764	84	35283	6.34	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	126.80%#
64) 1,2-Dichlorobenzene-d4	12.195	152	26816	5.97	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	119.40%
Target Compounds						Qvalue
3) Chloromethane	1.565	50	142994	14.94	ug/L	98
14) Carbon disulfide	2.867	76	29476	1.38	ug/L #	94
34) Trichloroethene	6.626	95	919100	144.13	ug/L	98
42) Toluene	8.012	91	1026	0.04	ug/L	86
47) Tetrachloroethene	8.587	164	6706	1.62	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011321\
Data File : VU042014.D
Acq On : 14 Jan 2021 05:04
Operator : SY/MD
Sample : M1129-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFT61

Quant Time: Jan 15 02:52:02 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR011321WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jan 15 02:46:12 2021
Response via : Initial Calibration

