

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011321\
 Data File : VU041981.D
 Acq On : 13 Jan 2021 15:27
 Operator : SY/MD
 Sample : VU0113WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK13

Quant Time: Jan 14 17:00:55 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR011321WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jan 14 17:00:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.259	114	134893	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	113497	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.809	152	50167	5.00	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	46466	4.56	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.20%
7) Chloroethane-d5	1.919	69	47220	4.90	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.00%
11) 1,1-Dichloroethene-d2	2.575	63	71878	3.51	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.20%
20) 2-Butanone-d5	4.655	46	160055	46.85	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.70%
24) Chloroform-d	5.076	84	89792	4.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.20%
26) 1,2-Dichloroethane-d4	5.713	65	55746	4.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.20%
32) Benzene-d6	5.739	84	189114	5.15	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.00%
36) 1,2-Dichloropropane-d6	6.700	67	58303	4.85	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.00%
41) Toluene-d8	7.902	98	147434	4.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.80%
43) trans-1,3-Dichloroprop...	8.182	79	23199	4.80	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.00%
46) 2-Hexanone-d5	8.639	63	124284	50.13	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.26%
57) 1,1,2,2-Tetrachloroeth...	10.754	84	42588	4.65	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.00%
64) 1,2-Dichlorobenzene-d4	12.188	152	46678	4.94	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.80%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011321\
Data File : VU041981.D
Acq On : 13 Jan 2021 15:27
Operator : SY/MD
Sample : VU0113WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK13

Quant Time: Jan 14 17:00:55 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR011321WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jan 14 17:00:00 2021
Response via : Initial Calibration

