

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011321\
 Data File : VU041999.D
 Acq On : 13 Jan 2021 23:04
 Operator : SY/MD
 Sample : M1065-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFXN0

Quant Time: Jan 15 02:47:19 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR011321WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jan 15 02:46:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.260	114	116283	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	111755	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.809	152	45076	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	43751	4.98	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	99.60%
7) Chloroethane-d5	1.919	69	42883	5.16	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.20%
11) 1,1-Dichloroethene-d2	2.578	63	93914	5.31	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	106.20%
20) 2-Butanone-d5	4.665	46	164410	55.83	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.66%
24) Chloroform-d	5.080	84	89517	4.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.60%
26) 1,2-Dichloroethane-d4	5.716	65	51602	5.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.40%
32) Benzene-d6	5.739	84	168843	4.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.40%
36) 1,2-Dichloropropane-d6	6.700	67	56722	4.79	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.80%
41) Toluene-d8	7.906	98	139935	4.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.60%
43) trans-1,3-Dichloroprop...	8.186	79	22751	4.78	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.60%
46) 2-Hexanone-d5	8.642	63	142878	58.52	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	117.04%
57) 1,1,2,2-Tetrachloroeth...	10.755	84	42928	4.76	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.20%
64) 1,2-Dichlorobenzene-d4	12.189	152	43850	5.17	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.40%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.385	85	15838	1.48	ug/L	99
9) Trichlorofluoromethane	2.144	101	8430	0.49	ug/L	97
10) 1,1,2-Trichloro-1,2,2-...	2.588	101	11836	1.27	ug/L	98
12) 1,1-Dichloroethene	2.588	96	50069	5.47	ug/L	99
13) Acetone	2.678	43	2987	1.62	ug/L	84
17) Methyl tert-butyl Ether	3.385	73	4641	0.19	ug/L #	94
18) trans-1,2-Dichloroethene	3.372	96	628	0.07	ug/L #	57
19) 1,1-Dichloroethane	3.887	63	28351	1.56	ug/L	97
22) cis-1,2-Dichloroethene	4.684	96	30374	3.14	ug/L	96
25) Chloroform	5.105	83	10958	0.60	ug/L	90
27) 1,2-Dichloroethane	5.806	62	3095	0.27	ug/L	98
29) 1,1,1-Trichloroethane	5.334	97	9604	0.59	ug/L	92
31) Carbon tetrachloride	5.536	117	4899	0.35	ug/L	87
33) Benzene	5.793	78	1821	0.04	ug/L	100
34) Trichloroethene	6.552	95	1134603	109.75	ug/L	97
47) Tetrachloroethene	8.559	164	604379	90.26	ug/L	95
51) Chlorobenzene	9.446	112	2376	0.10	ug/L	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011321\
Data File : VU041999.D
Acq On : 13 Jan 2021 23:04
Operator : SY/MD
Sample : M1065-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFXN0

Quant Time: Jan 15 02:47:19 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR011321WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jan 15 02:46:12 2021
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#)=qualifier out of range (m)=manual integration (+)=signals summed						

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011321\
Data File : VU041999.D
Acq On : 13 Jan 2021 23:04
Operator : SY/MD
Sample : M1065-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFXN0

Quant Time: Jan 15 02:47:19 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR011321WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jan 15 02:46:12 2021
Response via : Initial Calibration









#2
Dichlorodifluoromethane
Concen: 1.48 ug/L
RT: 1.385 min Scan# 14
Delta R.T. 0.000 min
Lab File: VU041999.D
Acq: 13 Jan 2021 23:04




Instrument :
MSVOA_U
ClientSampleId :
BFXN0


Tgt Ion: 85 Resp: 15838
Ion Ratio Lower Upper
85 100
87 32.8 25.7 38.5



#9
Trichlorofluoromethane
Concen: 0.49 ug/L
RT: 2.144 min Scan# 250
Delta R.T. 0.000 min
Lab File: VU041999.D
Acq: 13 Jan 2021 23:04

Tgt Ion:101 Resp: 8430
Ion Ratio Lower Upper
101 100
103 62.4 51.6 77.4









#10
1,1,2-Trichloro-1,2,2-trifluoroethane
Concen: 1.27 ug/L
RT: 2.588 min Scan# 388
Delta R.T. -0.003 min
Lab File: VU041999.D
Acq: 13 Jan 2021 23:04




Instrument :
MSVOA_U
ClientSampleId :
BFXN0


Tgt Ion:101 Resp: 11836
Ion Ratio Lower Upper
101 100
85 42.9 35.9 53.9
151 70.2 55.4 83.2



#12
1,1-Dichloroethene
Concen: 5.47 ug/L
RT: 2.588 min Scan# 388
Delta R.T. 0.000 min
Lab File: VU041999.D
Acq: 13 Jan 2021 23:04


Tgt Ion: 96 Resp: 50069
Ion Ratio Lower Upper
96 100
61 175.9 122.2 226.9
63 117.1 82.8 153.8








#13
Acetone
Concen: 1.62 ug/L
RT: 2.678 min Scan# 416
Delta R.T. 0.006 min
Lab File: VU041999.D
Acq: 13 Jan 2021 23:04

Instrument :
MSVOA_U
ClientSampleId :
BFXN0







Tgt Ion: 43 Resp: 2987
Ion Ratio Lower Upper
43 100
58 20.2 0.0 57.4



#17
Methyl tert-butyl Ether
Concen: 0.19 ug/L
RT: 3.385 min Scan# 636
Delta R.T. 0.006 min
Lab File: VU041999.D
Acq: 13 Jan 2021 23:04

Tgt Ion: 73 Resp: 4641
Ion Ratio Lower Upper
73 100
43 19.6 15.1 22.7
57 17.7 18.0 27.0#









#18
trans-1,2-Dichloroethene
Concen: 0.07 ug/L
RT: 3.372 min Scan# 632
Delta R.T. 0.006 min
Lab File: VU041999.D
Acq: 13 Jan 2021 23:04




Instrument :
MSVOA_U
ClientSampleId :
BFXN0


Tgt Ion: 96 Resp: 628
Ion Ratio Lower Upper
96 100
61 81.7 103.2 191.6#
98 45.8 43.8 81.3



#19
1,1-Dichloroethane
Concen: 1.56 ug/L
RT: 3.887 min Scan# 792
Delta R.T. 0.000 min
Lab File: VU041999.D
Acq: 13 Jan 2021 23:04

Tgt Ion: 63 Resp: 28351
Ion Ratio Lower Upper
63 100
65 34.3 22.7 42.1
83 14.9 9.4 17.6





#22

cis-1,2-Dichloroethene

Concen: 3.14 ug/L

RT: 4.684 min Scan# 1040

Delta R.T. 0.003 min





Lab File: VU041999.D

Acq: 13 Jan 2021 23:04

Tgt Ion: 96 Resp: 30374

Ion	Ratio	Lower	Upper
96	100		
61	138.0	100.3	186.3
68	0.0	0.0	0.0

Instrument :
MSVOA_U
ClientSampleId :
BFXN0



#25

Chloroform

Concen: 0.60 ug/L

RT: 5.105 min Scan# 1171




Delta R.T. 0.000 min


Lab File: VU041999.D

Acq: 13 Jan 2021 23:04

Tgt Ion: 83 Resp: 10958

Ion	Ratio	Lower	Upper
83	100		
85	73.5	46.1	85.7





#27

1,2-Dichloroethane

Concen: 0.27 ug/L

RT: 5.806 min Scan# 1389

Delta R.T. 0.000 min

Lab File: VU041999.D

Acq: 13 Jan 2021 23:04

Tgt Ion: 62 Resp: 3095

Ion Ratio Lower Upper

62 100





98 9.6 7.1 10.7

Instrument :

MSVOA_U

ClientSampleId :

BFXN0



#29

1,1,1-Trichloroethane

Concen: 0.59 ug/L

RT: 5.334 min Scan# 1242

Delta R.T. 0.003 min

Lab File: VU041999.D

Acq: 13 Jan 2021 23:04




Tgt Ion: 97 Resp: 9604


Ion Ratio Lower Upper

97 100

99 72.6 50.2 75.4

61 45.3 35.8 53.6









#31
Carbon tetrachloride
Concen: 0.35 ug/L
RT: 5.536 min Scan# 1305
Delta R.T. -0.003 min
Lab File: VU041999.D
Acq: 13 Jan 2021 23:04




Instrument :
MSVOA_U
ClientSampled :
BFXN0


Tgt Ion:117 Resp: 4899
Ion Ratio Lower Upper
117 100
119 83.2 76.6 114.8



#33
Benzene
Concen: 0.04 ug/L
RT: 5.793 min Scan# 1385
Delta R.T. 0.006 min
Lab File: VU041999.D
Acq: 13 Jan 2021 23:04

Tgt Ion: 78 Resp: 1821





#34

Trichloroethene

Concen: 109.75 ug/L

RT: 6.552 min Scan# 1621

Delta R.T. 0.000 min

Lab File: VU041999.D

Acq: 13 Jan 2021 23:04

Tgt Ion: 95 Resp: 1134603





Ion Ratio Lower Upper

95 100

97 63.6 43.6 81.0

132 87.4 64.3 119.3

130 91.9 66.7 123.9



#47

Tetrachloroethene

Concen: 90.26 ug/L

RT: 8.559 min Scan# 2245

Delta R.T. 0.003 min

Lab File: VU041999.D

Acq: 13 Jan 2021 23:04

Tgt Ion: 164 Resp: 604379




Ion Ratio Lower Upper


164 100

129 103.3 68.7 127.5

131 98.2 65.2 121.0


166 128.6 86.2 160.0





#51
Chlorobenzene
Concen: 0.10 ug/L
RT: 9.446 min Scan# 2521
Delta R.T. -0.003 min
Lab File: VU041999.D
Acq: 13 Jan 2021 23:04

Instrument :
MSVOA_U
ClientSampleId :
BFXN0



Tgt Ion:	112	Resp:	2376
Ion Ratio	Lower	Upper	
112	100		
114	31.3	22.3	41.3
77	66.3	48.2	72.4

