

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011323\
 Data File : VU052681.D
 Acq On : 13 Jan 2023 15:50
 Operator : JC/MD
 Sample : 01139-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH1Y8

Quant Time: Jan 13 22:27:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011023WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jan 11 00:53:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	148911	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	155608	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.806	152	72701	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.598	65	48674	3.721	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	74.400%
7) Chloroethane-d5	1.913	69	44325	4.159	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.200%
11) 1,1-Dichloroethene-d2	2.565	65	23537	4.085	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.800%
20) 2-Butanone-d5	4.639	46	166379	47.565	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.120%
24) Chloroform-d	5.057	84	109866	4.502	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.000%
26) 1,2-Dichloroethane-d4	5.697	65	60162	4.801	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.000%
32) Benzene-d6	5.723	84	219126	4.345	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.800%
36) 1,2-Dichloropropane-d6	6.684	67	74954	4.460	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.200%
41) Toluene-d8	7.893	98	185723	4.177	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.600%
43) trans-1,3-Dichloroprop...	8.176	79	27380	4.381	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.600%
46) 2-Hexanone-d5	8.630	63	131104	50.496	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.000%
56) 1,1,2,2-Tetrachloroeth...	10.752	84	56251	4.371	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.400%
66) 1,2-Dichlorobenzene-d4	12.189	152	67540	5.103	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.000%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.578	96	6122	0.474	ug/L #	1
14) Carbon disulfide	2.794	76	16603	0.372	ug/L	98
19) 1,1-Dichloroethane	3.868	63	20131	0.736	ug/L	97
21) 2-Butanone	4.729	43	8657	2.144	ug/L	85
22) cis-1,2-Dichloroethene	4.665	96	3044	0.215	ug/L #	80
34) Trichloroethene	6.540	95	204204	14.016	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011323\
Data File : VU052681.D
Acq On : 13 Jan 2023 15:50
Operator : JC/MD
Sample : 01139-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH1Y8

Quant Time: Jan 13 22:27:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011023WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jan 11 00:53:59 2023
Response via : Initial Calibration

