

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011325\
 Data File : VU062724.D
 Acq On : 13 Jan 2025 18:02
 Operator : MD/SY
 Sample : Q1058-13DL 5X
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH8Y2DL

Quant Time: Jan 13 19:04:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jan 13 18:58:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	96047	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	90498	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	43487	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	18990	3.570	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	71.400%	
7) Chloroethane-d5	1.901	69	18159	4.244	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	84.800%	
11) 1,1-Dichloroethene-d2	2.557	65	10407	3.776	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	75.600%	
20) 2-Butanone-d5	4.605	46	55642	51.179	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	102.360%	
24) Chloroform-d	5.049	84	58676	4.483	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	89.600%	
26) 1,2-Dichloroethane-d4	5.689	65	34611	4.876	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	97.600%	
32) Benzene-d6	5.714	84	97508	4.408	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	88.200%	
36) 1,2-Dichloropropane-d6	6.679	67	28694	4.678	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	93.600%	
41) Toluene-d8	7.888	98	88752	4.118	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	82.400%	
43) trans-1,3-Dichloroprop...	8.171	79	13707	4.483	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	89.600%	
46) 2-Hexanone-d5	8.621	63	46899	48.773	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	97.540%	
56) 1,1,2,2-Tetrachloroeth...	10.746	84	24388	4.748	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	95.000%	
66) 1,2-Dichlorobenzene-d4	12.187	152	34233	4.732	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	94.600%	
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	2.570	96	29701	4.981	ug/L	92
19) 1,1-Dichloroethane	3.856	63	28596	2.504	ug/L	95
34) Trichloroethene	6.534	95	3200	0.360	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011325\
 Data File : VU062724.D
 Acq On : 13 Jan 2025 18:02
 Operator : MD/SY
 Sample : Q1058-13DL 5X
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH8Y2DL

Quant Time: Jan 13 19:04:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Jan 13 18:58:58 2025
 Response via : Initial Calibration

