

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU011419\
 Data File : VU029158.D
 Acq On : 14 Jan 2019 12:11
 Operator : JC/SP
 Sample : K1136-05
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EB 2019.01.11

Quant Time: Jan 14 23:20:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 12 01:33:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	113004	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	198832	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	180764	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	72316	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	76009	54.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.82%	
35) Dibromofluoromethane	4.88	113	67603	52.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.90%	
50) Toluene-d8	7.56	98	248445	48.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.66%	
62) 4-Bromofluorobenzene	10.30	95	78487	42.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.84%	
Target Compounds						
16) Acetone	2.32	43	4478	5.088	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU011419\
Data File : VU029158.D
Acq On : 14 Jan 2019 12:11
Operator : JC/SP
Sample : K1136-05
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EB 2019.01.11

Quant Time: Jan 14 23:20:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 12 01:33:59 2019
Response via : Initial Calibration

