

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU011419\
 Data File : VU029159.D
 Acq On : 14 Jan 2019 12:36
 Operator : JC/SP
 Sample : K1135-06
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CPGP-6

Quant Time: Jan 16 05:51:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 12 01:33:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	111208	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	193647	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	177585	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	73321	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	74661	54.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.62%	
35) Dibromofluoromethane	4.88	113	65597	52.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.52%	
50) Toluene-d8	7.56	98	245280	49.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.98%	
62) 4-Bromofluorobenzene	10.30	95	77309	43.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.82%	
Target Compounds						
16) Acetone	2.32	43	3810	4.399	ug/l	96
64) Tetrachloroethene	8.23	164	1528	1.194	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU011419\
Data File : VU029159.D
Acq On : 14 Jan 2019 12:36
Operator : JC/SP
Sample : K1135-06
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CPGP-6

Quant Time: Jan 16 05:51:39 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 12 01:33:59 2019
Response via : Initial Calibration

