

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011421\
 Data File : VU042027.D
 Acq On : 14 Jan 2021 12:14
 Operator : SY/MD
 Sample : M1027-13DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFXK7DL

Quant Time: Jan 15 05:22:42 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR011321WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jan 15 05:20:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.260	114	124014	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	116605	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.809	152	46686	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	39745	4.25	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.00%
7) Chloroethane-d5	1.919	69	42536	4.80	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.00%
11) 1,1-Dichloroethene-d2	2.575	63	61610	3.27	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.40%
20) 2-Butanone-d5	4.655	46	145259	46.25	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.50%
24) Chloroform-d	5.080	84	87582	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.40%
26) 1,2-Dichloroethane-d4	5.716	65	50585	4.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.80%
32) Benzene-d6	5.742	84	166264	4.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.20%
36) 1,2-Dichloropropane-d6	6.700	67	55965	4.53	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.60%
41) Toluene-d8	7.903	98	154308	4.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.60%
43) trans-1,3-Dichloroprop...	8.186	79	23476	4.73	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.60%
46) 2-Hexanone-d5	8.642	63	130443	51.21	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.42%
57) 1,1,2,2-Tetrachloroeth...	10.755	84	42891	4.56	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.20%
64) 1,2-Dichlorobenzene-d4	12.189	152	44987	5.12	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.40%
Target Compounds						Qvalue
16) Methylene chloride	3.060	84	6027	0.53	ug/L	89
17) Methyl tert-butyl Ether	3.375	73	2691	0.11	ug/L #	77
22) cis-1,2-Dichloroethene	4.687	96	2177	0.21	ug/L	87
34) Trichloroethene	6.552	95	34526	3.20	ug/L	97
47) Tetrachloroethene	8.559	164	45092	6.45	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011421\
Data File : VU042027.D
Acq On : 14 Jan 2021 12:14
Operator : SY/MD
Sample : M1027-13DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFXK7DL

Quant Time: Jan 15 05:22:42 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR011321WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jan 15 05:20:45 2021
Response via : Initial Calibration

