

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011421\
 Data File : VU042030.D
 Acq On : 14 Jan 2021 13:27
 Operator : SY/MD
 Sample : M1027-15
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFXK9

Quant Time: Jan 15 05:23:52 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR011321WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jan 15 05:20:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.259	114	133022	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	123386	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.809	152	50331	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	40946	4.08	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.60%
7) Chloroethane-d5	1.919	69	41494	4.37	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.40%
11) 1,1-Dichloroethene-d2	2.575	63	61635	3.05	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.00%
20) 2-Butanone-d5	4.649	46	148711	44.15	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	88.30%
24) Chloroform-d	5.079	84	89869	4.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.40%
26) 1,2-Dichloroethane-d4	5.716	65	49977	4.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.20%
32) Benzene-d6	5.742	84	167571	4.20	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.00%
36) 1,2-Dichloropropane-d6	6.700	67	57885	4.43	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.60%
41) Toluene-d8	7.902	98	139276	3.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.00%
43) trans-1,3-Dichloroprop...	8.185	79	21897	4.17	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.40%
46) 2-Hexanone-d5	8.642	63	126855	47.06	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.12%
57) 1,1,2,2-Tetrachloroeth...	10.754	84	44655	4.48	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.60%
64) 1,2-Dichlorobenzene-d4	12.188	152	44762	4.72	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.40%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.385	85	6485	0.53	ug/L	97
3) Chloromethane	1.523	50	1599	0.10	ug/L	91
13) Acetone	2.658	43	4357	2.07	ug/L	87
16) Methylene chloride	3.060	84	1742	0.14	ug/L	89
17) Methyl tert-butyl Ether	3.379	73	22146	0.81	ug/L	98
22) cis-1,2-Dichloroethene	4.681	96	8322	0.75	ug/L	83
25) Chloroform	5.102	83	2799	0.13	ug/L	90
27) 1,2-Dichloroethane	5.809	62	1529	0.11	ug/L #	83
29) 1,1,1-Trichloroethane	5.330	97	1797	0.10	ug/L	95
34) Trichloroethene	6.552	95	183717	16.10	ug/L	96
47) Tetrachloroethene	8.558	164	132366	17.90	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011421\
Data File : VU042030.D
Acq On : 14 Jan 2021 13:27
Operator : SY/MD
Sample : M1027-15
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFXK9

Quant Time: Jan 15 05:23:52 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR011321WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jan 15 05:20:45 2021
Response via : Initial Calibration

