

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011421\
 Data File : VU042038.D
 Acq On : 14 Jan 2021 17:09
 Operator : SY/MD
 Sample : VU0114WBL02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK22

Quant Time: Jan 15 06:33:14 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR011321WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jan 15 06:32:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.260	114	115474	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	99826	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.809	152	42852	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	39371	4.52	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.40%
7) Chloroethane-d5	1.919	69	41886	5.08	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.60%
11) 1,1-Dichloroethene-d2	2.575	63	59263	3.38	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.60%
20) 2-Butanone-d5	4.652	46	162546	55.59	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.18%
24) Chloroform-d	5.076	84	89585	5.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.40%
26) 1,2-Dichloroethane-d4	5.713	65	51818	5.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.60%
32) Benzene-d6	5.739	84	173808	5.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.60%
36) 1,2-Dichloropropane-d6	6.700	67	63419	5.99	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	119.80%
41) Toluene-d8	7.903	98	154209	5.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.80%
43) trans-1,3-Dichloroprop...	8.182	79	23093	5.43	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	108.60%
46) 2-Hexanone-d5	8.639	63	121733	55.82	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	111.64%
57) 1,1,2,2-Tetrachloroeth...	10.755	84	43816	5.44	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	108.80%
64) 1,2-Dichlorobenzene-d4	12.189	152	46337	5.74	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.80%
Target Compounds						Qvalue
3) Chloromethane	1.523	50	1662	0.12	ug/L	# 81
16) Methylene chloride	3.054	84	1577	0.15	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011421\
Data File : VU042038.D
Acq On : 14 Jan 2021 17:09
Operator : SY/MD
Sample : VU0114WBL02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK22

Quant Time: Jan 15 06:33:14 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR011321WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jan 15 06:32:40 2021
Response via : Initial Calibration

