

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011425\
 Data File : VU062731.D
 Acq On : 14 Jan 2025 09:59
 Operator : MD/SY
 Sample : VU0114WBL01
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK139

Quant Time: Jan 15 00:31:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jan 15 00:31:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	91049	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	88045	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	39407	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	19569	3.881	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.600%
7) Chloroethane-d5	1.904	69	18059	4.452	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.000%
11) 1,1-Dichloroethene-d2	2.557	65	10629	4.069	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.400%
20) 2-Butanone-d5	4.618	46	58576	56.836	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.680%
24) Chloroform-d	5.049	84	61504	4.957	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.200%
26) 1,2-Dichloroethane-d4	5.689	65	35142	5.222	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.400%
32) Benzene-d6	5.714	84	101501	4.716	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.400%
36) 1,2-Dichloropropane-d6	6.679	67	30594	5.127	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.600%
41) Toluene-d8	7.888	98	93304	4.450	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.000%
43) trans-1,3-Dichloroprop...	8.171	79	14101	4.741	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.800%
46) 2-Hexanone-d5	8.624	63	48870	52.239	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.480%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	26671	5.337	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.800%
66) 1,2-Dichlorobenzene-d4	12.184	152	35210	5.371	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011425\
Data File : VU062731.D
Acq On : 14 Jan 2025 09:59
Operator : MD/SY
Sample : VU0114WBL01
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK139

Quant Time: Jan 15 00:31:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jan 15 00:31:11 2025
Response via : Initial Calibration

