

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011425\
 Data File : VU062732.D
 Acq On : 14 Jan 2025 10:22
 Operator : MD/SY
 Sample : Q1058-02
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH8X1

Quant Time: Jan 15 00:32:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jan 15 00:31:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	94037	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	88105	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	42040	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	17904	3.438	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	68.800%
7) Chloroethane-d5	1.904	69	17249	4.117	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.400%
11) 1,1-Dichloroethene-d2	2.557	65	10025	3.716	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.400%
20) 2-Butanone-d5	4.608	46	49499	46.502	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.000%
24) Chloroform-d	5.049	84	56302	4.394	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.800%
26) 1,2-Dichloroethane-d4	5.692	65	32556	4.684	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.600%
32) Benzene-d6	5.717	84	94143	4.371	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.400%
36) 1,2-Dichloropropane-d6	6.679	67	28049	4.697	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.000%
41) Toluene-d8	7.888	98	86052	4.101	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.000%
43) trans-1,3-Dichloroprop...	8.174	79	12514	4.204	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.000%
46) 2-Hexanone-d5	8.621	63	38823	41.471	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.940%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	23510	4.701	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.000%
66) 1,2-Dichlorobenzene-d4	12.183	152	33047	4.725	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.400%
Target Compounds	Qvalue					
12) 1,1-Dichloroethene	2.573	96	8442	1.446	ug/L #	63
17) Methyl tert-butyl Ether	3.348	73	11097	0.692	ug/L	98
19) 1,1-Dichloroethane	3.856	63	23927	2.140	ug/L	93
22) cis-1,2-Dichloroethene	4.656	96	9204	1.351	ug/L	95
29) 1,1,1-Trichloroethane	5.309	97	5255	0.393	ug/L	92
34) Trichloroethene	6.531	95	61620	7.126	ug/L	98
42) Toluene	7.962	91	8627	0.311	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011425\
Data File : VU062732.D
Acq On : 14 Jan 2025 10:22
Operator : MD/SY
Sample : Q1058-02
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH8X1

Quant Time: Jan 15 00:32:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jan 15 00:31:11 2025
Response via : Initial Calibration

