

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011425\
 Data File : VU062751.D
 Acq On : 14 Jan 2025 18:10
 Operator : MD/SY
 Sample : VIBLK196
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK196

Quant Time: Jan 15 00:37:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jan 15 00:31:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.235	114	188716	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.405	117	178821	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.801	152	90681	5.000	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	34708	3.321	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	66.400%
7) Chloroethane-d5	1.901	69	32399	3.853	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	77.000%
11) 1,1-Dichloroethene-d2	2.554	65	20312	3.751	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.000%
20) 2-Butanone-d5	4.615	46	97849	45.806	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	91.620%
24) Chloroform-d	5.049	84	115177	4.479	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.600%
26) 1,2-Dichloroethane-d4	5.689	65	66527	4.770	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.400%
32) Benzene-d6	5.714	84	189117	4.326	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.600%
36) 1,2-Dichloropropane-d6	6.676	67	55399	4.571	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.400%
41) Toluene-d8	7.888	98	177453	4.167	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.400%
43) trans-1,3-Dichloroprop...	8.171	79	26284	4.351	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.000%
46) 2-Hexanone-d5	8.624	63	99926	52.592	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.180%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	49684	4.895	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.000%
66) 1,2-Dichlorobenzene-d4	12.184	152	70994	4.706	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011425\
Data File : VU062751.D
Acq On : 14 Jan 2025 18:10
Operator : MD/SY
Sample : VIBLK196
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK196

Quant Time: Jan 15 00:37:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jan 15 00:31:11 2025
Response via : Initial Calibration

