

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU011519\
 Data File : VU029175.D
 Acq On : 15 Jan 2019 11:05
 Operator : JC/SP
 Sample : K1135-01ME
 Misc : 5.08g/10mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CPSB-19(10-15)ME

Quant Time: Jan 16 05:40:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 12 01:33:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.97	168	109339	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	193045	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	189818	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	90798	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.30	65	72726	53.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.62%	
35) Dibromofluoromethane	4.88	113	64684	51.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.38%	
50) Toluene-d8	7.56	98	245752	49.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.48%	
62) 4-Bromofluorobenzene	10.31	95	92864	52.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.60%	
Target Compounds						
65) Chlorobenzene	9.12	112	15854	3.726	ug/l	95
78) n-propylbenzene	10.59	91	7922	1.000	ug/l	98
85) sec-Butylbenzene	11.32	105	9565	1.430	ug/l	93
89) n-Butylbenzene	11.89	91	7794	1.532	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU011519\
Data File : VU029175.D
Acq On : 15 Jan 2019 11:05
Operator : JC/SP
Sample : K1135-01ME
Misc : 5.08g/10mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CPSB-19(10-15)ME

Quant Time: Jan 16 05:40:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 12 01:33:59 2019
Response via : Initial Calibration

