

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU011519\
 Data File : VU029181.D
 Acq On : 15 Jan 2019 13:30
 Operator : JC/SP
 Sample : PB116371ZHE#03
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB116371ZHE#03

Quant Time: Jan 16 06:05:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 12 01:33:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	114337	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	202507	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	189158	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	76827	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	77846	55.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.16%	
35) Dibromofluoromethane	4.88	113	67002	51.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.08%	
50) Toluene-d8	7.56	98	260587	50.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.56%	
62) 4-Bromofluorobenzene	10.30	95	82349	44.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.44%	
Target Compounds						
16) Acetone	2.32	43	4785	5.374	ug/l	98
20) Methylene Chloride	2.70	84	5759	3.809	ug/l	97
43) Isopropyl Acetate	5.55	43	8885	2.607	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU011519\
Data File : VU029181.D
Acq On : 15 Jan 2019 13:30
Operator : JC/SP
Sample : PB116371ZHE#03
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB116371ZHE#03

Quant Time: Jan 16 06:05:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 12 01:33:59 2019
Response via : Initial Calibration

