

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011524\
 Data File : VU057532.D
 Acq On : 15 Jan 2024 16:12
 Operator : MD/SY
 Sample : P1120-20
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGWJ6

Quant Time: Jan 16 00:36:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011124WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jan 16 00:31:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.241	114	256170	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	263145	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	98040	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	130440	5.019	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.400%
7) Chloroethane-d5	1.914	69	111876	5.532	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	110.600%
11) 1,1-Dichloroethene-d2	2.563	65	55363	5.037	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	100.800%
20) 2-Butanone-d5	4.640	46	236935	48.195	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.380%
24) Chloroform-d	5.055	84	246828	5.497	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.000%
26) 1,2-Dichloroethane-d4	5.695	65	122482	5.663	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	113.200%
32) Benzene-d6	5.721	84	469659	4.893	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.800%
36) 1,2-Dichloropropane-d6	6.685	67	152524	5.097	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.000%
41) Toluene-d8	7.891	98	284855	3.516	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	70.400%
43) trans-1,3-Dichloroprop...	8.177	79	35592	3.421	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	68.400%
46) 2-Hexanone-d5	8.630	63	133572	35.461	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	70.920%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	97935	4.845	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	111878	5.682	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.600%
Target Compounds						Qvalue
5) Vinyl chloride	1.602	62	83514	4.321	ug/L	94
8) Chloroethane	1.930	64	4017	0.333	ug/L #	69
12) 1,1-Dichloroethene	2.576	96	662139	44.240	ug/L #	64
14) Carbon disulfide	2.791	76	32518	0.778	ug/L	99
15) Methyl Acetate	2.962	43	5693	0.852	ug/L	90
16) Methylene chloride	3.042	84	3081	0.152	ug/L	87
18) trans-1,2-Dichloroethene	3.357	96	1673	0.113	ug/L	88
19) 1,1-Dichloroethane	3.862	63	333966	9.963	ug/L	98
22) cis-1,2-Dichloroethene	4.660	96	446138	25.797	ug/L	97
27) 1,2-Dichloroethane	5.791	62	7058	0.337	ug/L #	93
34) Trichloroethene	6.537	95	300019	14.772	ug/L	98
37) 1,2-Dichloropropane	6.785	63	7792	0.355	ug/L	100
42) Toluene	7.965	91	10517	0.138	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011524\
Data File : VU057532.D
Acq On : 15 Jan 2024 16:12
Operator : MD/SY
Sample : P1120-20
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGWJ6

Quant Time: Jan 16 00:36:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011124WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jan 16 00:31:32 2024
Response via : Initial Calibration

