

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011525\
 Data File : VU062790.D
 Acq On : 16 Jan 2025 01:38
 Operator : MD/SY
 Sample : P1071-18
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :

Quant Time: Jan 16 02:10:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jan 16 01:55:19 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	95272	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	89391	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	41470	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	15737	2.982	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	59.600%	
7) Chloroethane-d5	1.904	69	16275	3.834	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	76.600%	
11) 1,1-Dichloroethene-d2	2.557	65	9629	3.523	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	70.400%	
20) 2-Butanone-d5	4.602	46	65486	60.724	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	121.440%	
24) Chloroform-d	5.049	84	62292	4.798	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.000%	
26) 1,2-Dichloroethane-d4	5.689	65	36791	5.225	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	104.400%	
32) Benzene-d6	5.714	84	97195	4.448	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	89.000%	
36) 1,2-Dichloropropane-d6	6.679	67	29963	4.946	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	99.000%	
41) Toluene-d8	7.888	98	88092	4.138	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	82.800%	
43) trans-1,3-Dichloroprop...	8.174	79	14207	4.704	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	94.000%	
46) 2-Hexanone-d5	8.621	63	51568	54.293	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	108.580%	
56) 1,1,2,2-Tetrachloroeth...	10.746	84	28135	5.545	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	111.000%	
66) 1,2-Dichlorobenzene-d4	12.184	152	36621	5.308	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	106.200%	
Target Compounds						
					Qvalue	
5) Vinyl chloride	1.599	62	1201	0.175	ug/L #	5
12) 1,1-Dichloroethene	2.570	96	62417	10.554	ug/L #	82
18) trans-1,2-Dichloroethene	3.341	96	3304	0.521	ug/L	80
19) 1,1-Dichloroethane	3.856	63	81831	7.223	ug/L	95
22) cis-1,2-Dichloroethene	4.650	96	116874	16.938	ug/L	99
27) 1,2-Dichloroethane	5.792	62	1812	0.199	ug/L #	79
34) Trichloroethene	6.534	95	13297	1.516	ug/L	95
35) Methylcyclohexane	6.679	83	8758	0.811	ug/L #	21
37) 1,2-Dichloropropane	6.676	63	3668	0.605	ug/L #	85
61) 1,2,3-Trichloropropane	11.801	75	4123	1.121	ug/L #	61


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011525\
Data File : VU062790.D
Acq On : 16 Jan 2025 01:38
Operator : MD/SY
Sample : P1071-18
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :

Quant Time: Jan 16 02:10:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jan 16 01:55:19 2025
Response via : Initial Calibration









#5
Vinyl chloride
Concen: 0.175 ug/L
RT: 1.599 min Scan# 90
Delta R.T. 0.000 min
Lab File: VU062790.D
Acq: 16 Jan 2025 01:38




Instrument :
MSVOA_U
ClientSampleId :


Tgt Ion: 62 Resp: 1201
Ion Ratio Lower Upper
62 100
64 87.4 23.4 43.4#



#12
1,1-Dichloroethene
Concen: 10.554 ug/L
RT: 2.570 min Scan# 398
Delta R.T. 0.003 min
Lab File: VU062790.D
Acq: 16 Jan 2025 01:38


Tgt Ion: 96 Resp: 62417
Ion Ratio Lower Upper
96 100
61 163.9 120.4 223.6
63 78.9 82.3 152.9#








#18
trans-1,2-Dichloroethene
Concen: 0.521 ug/L
RT: 3.341 min Scan# 61
Delta R.T. 0.000 min
Lab File: VU062790.D
Acq: 16 Jan 2025 01:38

Instrument :
MSVOA_U
ClientSampleId :




Tgt Ion: 96 Resp: 3304

Ion	Ratio	Lower	Upper
96	100		
61	163.5	94.4	175.2
98	68.8	42.8	79.4






#19
1,1-Dichloroethane
Concen: 7.223 ug/L
RT: 3.856 min Scan# 798
Delta R.T. 0.003 min
Lab File: VU062790.D
Acq: 16 Jan 2025 01:38



Tgt Ion: 63 Resp: 81831

Ion	Ratio	Lower	Upper
63	100		
65	32.8	20.7	38.5
83	16.1	10.5	19.5









#22
 cis-1,2-Dichloroethene
 Concen: 16.938 ug/L
 RT: 4.650 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VU062790.D
 Acq: 16 Jan 2025 01:38




Instrument :
 MSVOA_U
 ClientSampleId :


Tgt Ion: 96 Resp: 116874
 Ion Ratio Lower Upper
 96 100
 61 120.4 84.8 157.6
 98 63.5 43.1 80.1



#27
 1,2-Dichloroethane
 Concen: 0.199 ug/L
 RT: 5.792 min Scan# 1400
 Delta R.T. 0.013 min
 Lab File: VU062790.D
 Acq: 16 Jan 2025 01:38

Tgt Ion: 62 Resp: 1812
 Ion Ratio Lower Upper
 62 100
 98 0.0 5.8 8.6#









#34
Trichloroethene
Concen: 1.516 ug/L
RT: 6.534 min Scan# 1630
Delta R.T. 0.003 min
Lab File: VU062790.D
Acq: 16 Jan 2025 01:38




Instrument :
MSVOA_U
ClientSampleId :


Tgt Ion: 95 Resp: 13297
Ion Ratio Lower Upper
95 100
97 61.7 47.5 88.3
132 95.0 66.9 124.3
130 106.5 68.6 127.4



#35
Methylcyclohexane
Concen: 0.811 ug/L
RT: 6.679 min Scan# 1676
Delta R.T. -0.071 min
Lab File: VU062790.D
Acq: 16 Jan 2025 01:38

Tgt Ion: 83 Resp: 8758
Ion Ratio Lower Upper
83 100
55 0.0 57.2 85.8#
98 0.5 36.6 55.0#





#37

1,2-Dichloropropane

Concen: 0.605 ug/L

RT: 6.676 min Scan# 1

Delta R.T. -0.103 min

Lab File: VU062790.D

Acq: 16 Jan 2025 01:38

Instrument :

MSVOA_U





ClientSampleId :

Tgt Ion: 63 Resp: 3668

Ion Ratio Lower Upper

63 100

112 0.0 3.8 5.8#



#61

1,2,3-Trichloropropane

Concen: 1.121 ug/L

RT: 11.801 min Scan# 3269

Delta R.T. 0.990 min

Lab File: VU062790.D

Acq: 16 Jan 2025 01:38

Tgt Ion: 75 Resp: 4123

Ion Ratio Lower Upper

75 100

110 1.0 33.2 49.8#

77 35.6 25.9 38.9

