

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011525\
 Data File : VU062789.D
 Acq On : 16 Jan 2025 01:14
 Operator : MD/SY
 Sample : Q1071-17
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GCQ81

Quant Time: Jan 16 02:10:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jan 16 01:55:19 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	95715	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	88467	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	43172	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	14955	2.821	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	56.400%
7) Chloroethane-d5	1.901	69	15135	3.549	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	71.000%
11) 1,1-Dichloroethene-d2	2.557	65	8961	3.263	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.200%
20) 2-Butanone-d5	4.605	46	61523	56.785	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.560%
24) Chloroform-d	5.049	84	59321	4.548	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.000%
26) 1,2-Dichloroethane-d4	5.689	65	34113	4.822	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.400%
32) Benzene-d6	5.714	84	93779	4.337	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.800%
36) 1,2-Dichloropropane-d6	6.679	67	28441	4.743	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.800%
41) Toluene-d8	7.888	98	83242	3.951	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.000%
43) trans-1,3-Dichloroprop...	8.171	79	14297	4.784	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.600%
46) 2-Hexanone-d5	8.621	63	49426	52.581	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.160%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	26404	5.258	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.200%
66) 1,2-Dichlorobenzene-d4	12.183	152	34101	4.748	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.000%
Target Compounds						
12) 1,1-Dichloroethene	2.570	96	26430	4.448	ug/L	95
17) Methyl tert-butyl Ether	3.348	73	98930	6.058	ug/L	100
19) 1,1-Dichloroethane	3.856	63	74665	6.560	ug/L	97
22) cis-1,2-Dichloroethene	4.653	96	44938	6.482	ug/L	96
34) Trichloroethene	6.534	95	3697	0.426	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011525\
Data File : VU062789.D
Acq On : 16 Jan 2025 01:14
Operator : MD/SY
Sample : Q1071-17
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GCQ81

Quant Time: Jan 16 02:10:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR010225WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jan 16 01:55:19 2025
Response via : Initial Calibration

