

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU011619\
 Data File : VU029196.D
 Acq On : 16 Jan 2019 09:43
 Operator : JC/SP
 Sample : VU0116WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0116WBL01

Quant Time: Jan 17 01:06:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 12 01:33:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	102786	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	180941	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	169470	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	65631	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	71035	55.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	111.82%	
35) Dibromofluoromethane	4.88	113	62424	53.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.44%	
50) Toluene-d8	7.56	98	231781	50.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.12%	
62) 4-Bromofluorobenzene	10.30	95	71784	43.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.28%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU011619\
Data File : VU029196.D
Acq On : 16 Jan 2019 09:43
Operator : JC/SP
Sample : VU0116WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0116WBL01

Quant Time: Jan 17 01:06:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 12 01:33:59 2019
Response via : Initial Calibration

