

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU011619\
 Data File : VU029205.D
 Acq On : 16 Jan 2019 13:29
 Operator : JC/SP
 Sample : K1147-21
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TP-7

Quant Time: Jan 17 01:25:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 12 01:33:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	108088	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	190404	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	185059	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	74555	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	75350	56.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.78%	
35) Dibromofluoromethane	4.88	113	66651	54.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.00%	
50) Toluene-d8	7.56	98	255210	52.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.76%	
62) 4-Bromofluorobenzene	10.31	95	78953	45.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.18%	
Target Compounds						
16) Acetone	2.32	43	4988	5.926	ug/l	99
20) Methylene Chloride	2.70	84	131266	91.850	ug/l	99
43) Isopropyl Acetate	5.55	43	7473	2.332	ug/l	94
84) 1,2,4-Trimethylbenzene	11.15	105	4752	1.033	ug/l	98
95) Naphthalene	13.75	128	8122	1.408	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU011619\
Data File : VU029205.D
Acq On : 16 Jan 2019 13:29
Operator : JC/SP
Sample : K1147-21
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TP-7

Quant Time: Jan 17 01:25:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 12 01:33:59 2019
Response via : Initial Calibration

