

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU011619\
 Data File : VU029208.D
 Acq On : 16 Jan 2019 14:42
 Operator : JC/SP
 Sample : K1012-08
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EO-01-011518-A

Quant Time: Jan 17 01:32:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 12 01:33:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	108963	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	194109	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	185445	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	77939	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	75243	55.86	ug/l	0.00
Spiked Amount			Recovery	=	111.72%	
35) Dibromofluoromethane	4.88	113	66469	52.83	ug/l	0.00
Spiked Amount			Recovery	=	105.66%	
50) Toluene-d8	7.56	98	255040	51.34	ug/l	0.00
Spiked Amount			Recovery	=	102.68%	
62) 4-Bromofluorobenzene	10.30	95	81273	45.53	ug/l	0.00
Spiked Amount			Recovery	=	91.06%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.32	43	11170	13.164	ug/l	93
20) Methylene Chloride	2.70	84	359121	249.267	ug/l	98
25) 2-Butanone	4.29	43	2610	2.277	ug/l	93
43) Isopropyl Acetate	5.55	43	10071	3.083	ug/l	95
68) m/p-Xylenes	9.37	106	3345	1.240	ug/l	91
84) 1,2,4-Trimethylbenzene	11.15	105	8186	1.702	ug/l	99
95) Naphthalene	13.75	128	22247	3.689	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU011619\
Data File : VU029208.D
Acq On : 16 Jan 2019 14:42
Operator : JC/SP
Sample : K1012-08
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EO-01-011518-A

Quant Time: Jan 17 01:32:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U010319W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 12 01:33:59 2019
Response via : Initial Calibration

