

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011623\
 Data File : VU052709.D
 Acq On : 16 Jan 2023 18:31
 Operator : JC/MD
 Sample : VIBLK143
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK143

Quant Time: Jan 17 05:35:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011023WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jan 11 00:53:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.246	114	123052	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	123159	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.806	152	54474	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	32332	2.991	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	59.800%
7) Chloroethane-d5	1.915	69	33036	3.751	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	75.000%
11) 1,1-Dichloroethene-d2	2.568	65	18004	3.782	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.600%
20) 2-Butanone-d5	4.642	46	152317	52.695	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.400%
24) Chloroform-d	5.060	84	97918	4.856	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.200%
26) 1,2-Dichloroethane-d4	5.697	65	52243	5.045	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.800%
32) Benzene-d6	5.722	84	179973	4.509	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.200%
36) 1,2-Dichloropropane-d6	6.687	67	61184	4.600	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.000%
41) Toluene-d8	7.893	98	139963	3.977	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.600%
43) trans-1,3-Dichloroprop...	8.176	79	21543	4.355	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.000%
46) 2-Hexanone-d5	8.632	63	105019	51.106	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.220%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	51472	5.054	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.000%
66) 1,2-Dichlorobenzene-d4	12.188	152	54236	5.469	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.400%
Target Compounds						Qvalue
16) Methylene chloride	3.047	84	4853	0.342	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU011623\
Data File : VU052709.D
Acq On : 16 Jan 2023 18:31
Operator : JC/MD
Sample : VIBLK143
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK143

Quant Time: Jan 17 05:35:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR011023WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jan 11 00:53:59 2023
Response via : Initial Calibration

